

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
 Data File : VY012936.D
 Acq On : 14 Mar 2023 12:39
 Operator : KP/MD
 Sample : 01749-17
 Misc : 6.73g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CF-19

Quant Time: Mar 15 01:41:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	80861	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	134865	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	119936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	48206	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	52743	65.695	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.400%
35) Dibromofluoromethane	7.716	113	46597	57.251	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.500%
50) Toluene-d8	10.179	98	161491	54.274	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.540%
62) 4-Bromofluorobenzene	12.483	95	44566	44.375	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.760%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
Data File : VY012936.D
Acq On : 14 Mar 2023 12:39
Operator : KP/MD
Sample : 01749-17
Misc : 6.73g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CF-19

Quant Time: Mar 15 01:41:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY012936.D
Acq: 14 Mar 2023 12:39




Instrument :
MSVOA_Y
ClientSampleId :
CF-19


Tgt Ion:168 Resp: 80861
Ion Ratio Lower Upper
168 100
99 52.7 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 65.695 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.001 min
Lab File: VY012936.D
Acq: 14 Mar 2023 12:39

Tgt Ion: 65 Resp: 52743
Ion Ratio Lower Upper
65 100
67 51.4 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY012936.D

Acq: 14 Mar 2023 12:39

Instrument :

MSVOA_Y

ClientSampleId :

CF-19





Tgt Ion:114 Resp: 134865

Ion Ratio Lower Upper

114 100

63 19.6 0.0 37.4

88 14.7 0.0 28.6



#35

Dibromofluoromethane

Concen: 57.251 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY012936.D

Acq: 14 Mar 2023 12:39




Tgt Ion:113 Resp: 46597


Ion Ratio Lower Upper

113 100

111 100.5 82.2 123.2

192 20.5 16.8 25.2





#50

Toluene-d8

Concen: 54.274 ug/l

RT: 10.179 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY012936.D

Acq: 14 Mar 2023 12:39

Instrument :

MSVOA_Y

ClientSampleId :





CF-19

Tgt Ion: 98 Resp: 161491

Ion Ratio Lower Upper

98 100

100 63.3 50.9 76.3



#62

4-Bromofluorobenzene

Concen: 44.375 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VY012936.D

Acq: 14 Mar 2023 12:39




Tgt Ion: 95 Resp: 44566


Ion Ratio Lower Upper

95 100

174 92.6 0.0 176.6

176 88.3 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY012936.D
Acq: 14 Mar 2023 12:39

Instrument :
MSVOA_Y
ClientSampleId :
CF-19

Tgt Ion:117 Resp: 119936
Ion Ratio Lower Upper
117 100
82 55.9 45.4 68.0
119 32.4 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY012936.D
Acq: 14 Mar 2023 12:39

Tgt Ion:152 Resp: 48206
Ion Ratio Lower Upper
152 100
115 54.4 27.1 81.3
150 155.8 0.0 347.8

