

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
 Data File : VY012942.D
 Acq On : 14 Mar 2023 15:00
 Operator : KP/MD
 Sample : 01874-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-209

Quant Time: Mar 15 01:43:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	75874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	129070	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	120514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	54626	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	48869	64.871	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.740%
35) Dibromofluoromethane	7.716	113	23483	30.148	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	60.300%
50) Toluene-d8	10.179	98	158262	55.577	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.160%
62) 4-Bromofluorobenzene	12.483	95	45989	47.848	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.700%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031423\
Data File : VY012942.D
Acq On : 14 Mar 2023 15:00
Operator : KP/MD
Sample : 01874-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-209

Quant Time: Mar 15 01:43:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY012942.D
Acq: 14 Mar 2023 15:00




Instrument :
MSVOA_Y
ClientSampleId :
VNJ-209


Tgt Ion:168 Resp: 75874
Ion Ratio Lower Upper
168 100
99 52.8 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 64.871 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. -0.000 min
Lab File: VY012942.D
Acq: 14 Mar 2023 15:00

Tgt Ion: 65 Resp: 48869
Ion Ratio Lower Upper
65 100
67 52.6 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY012942.D

Acq: 14 Mar 2023 15:00

Instrument :

MSVOA_Y

ClientSampleId :

VNJ-209





Tgt Ion:114 Resp: 129070

Ion Ratio Lower Upper

114 100

63 18.6 0.0 37.4

88 14.8 0.0 28.6



#35

Dibromofluoromethane

Concen: 30.148 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY012942.D

Acq: 14 Mar 2023 15:00




Tgt Ion:113 Resp: 23483

Ion Ratio Lower Upper

113 100

111 104.5 82.2 123.2

192 22.4 16.8 25.2








#50
Toluene-d8
Concen: 55.577 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY012942.D
Acq: 14 Mar 2023 15:00




Instrument :
MSVOA_Y
ClientSampleId :
VNJ-209


Tgt Ion: 98 Resp: 158262
Ion Ratio Lower Upper
98 100
100 62.8 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 47.848 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY012942.D
Acq: 14 Mar 2023 15:00

Tgt Ion: 95 Resp: 45989
Ion Ratio Lower Upper
95 100
174 93.7 0.0 176.6
176 89.5 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY012942.D
Acq: 14 Mar 2023 15:00

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-209

Tgt Ion:117 Resp: 120514
Ion Ratio Lower Upper
117 100
82 54.6 45.4 68.0
119 32.2 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY012942.D
Acq: 14 Mar 2023 15:00

Tgt Ion:152 Resp: 54626
Ion Ratio Lower Upper
152 100
115 54.8 27.1 81.3
150 155.0 0.0 347.8

