

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031523\
 Data File : VY012951.D
 Acq On : 15 Mar 2023 10:42
 Operator : KP/MD
 Sample : VY0315SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0315SBL01

Quant Time: Mar 16 00:31:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	83281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	141006	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	130934	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	57323	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	53027	64.130	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.260%
35) Dibromofluoromethane	7.722	113	48290	56.747	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.500%
50) Toluene-d8	10.185	98	169746	54.564	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.120%
62) 4-Bromofluorobenzene	12.483	95	50588	48.178	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031523\
Data File : VY012951.D
Acq On : 15 Mar 2023 10:42
Operator : KP/MD
Sample : VY0315SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0315SBL01

Quant Time: Mar 16 00:31:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

