

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031523\
 Data File : VY012956.D
 Acq On : 15 Mar 2023 15:45
 Operator : KP/MD
 Sample : 01936-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-03152023

Quant Time: Mar 16 00:19:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1727	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	3042	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	2889	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	724	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	1650	96.227	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 192.460%#	
35) Dibromofluoromethane	7.716	113	1239	67.489	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 134.980%	
50) Toluene-d8	10.185	98	3501	52.164	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.320%	
62) 4-Bromofluorobenzene	12.489	95	829	36.596	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 73.200%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031523\
Data File : VY012956.D
Acq On : 15 Mar 2023 15:45
Operator : KP/MD
Sample : 01936-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-03152023

Quant Time: Mar 16 00:19:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

