



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031523\
Data File : VY012949.D
Acq On : 15 Mar 2023 09:12
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Title : SW846 8260
Last Update : Fri Mar 03 01:13:13 2023



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	9370	PASS
75	95	30	60	52.4	23883	PASS
95	95	100	100	100.0	45541	PASS
96	95	5	9	6.8	3092	PASS
173	174	0.00	2	1.9	801	PASS
174	95	50	100	93.5	42589	PASS
175	174	5	9	7.4	3153	PASS
176	174	95	101	96.7	41179	PASS
177	176	5	9	6.6	2704	PASS