

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
 Data File : VY017668.D
 Acq On : 15 Mar 2024 13:25
 Operator : SY/MD
 Sample : P1767-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-031424

Quant Time: Mar 16 00:15:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	24967	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	31480	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	20344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4885	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	6628	28.011	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	56.020%
35) Dibromofluoromethane	7.709	113	9192	41.903	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.800%
50) Toluene-d8	10.179	98	33143	43.213	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	86.420%
62) 4-Bromofluorobenzene	12.483	95	5348	22.187	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	44.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
Data File : VY017668.D
Acq On : 15 Mar 2024 13:25
Operator : SY/MD
Sample : P1767-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-031424

Quant Time: Mar 16 00:15:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

