

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
 Data File : VY017670.D
 Acq On : 15 Mar 2024 14:12
 Operator : SY/MD
 Sample : P1746-03
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-4.0-6.0-DUP

Quant Time: Mar 16 00:16:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	119368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	217629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	192770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	56833	50.237	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.480%
35) Dibromofluoromethane	7.716	113	70285	46.347	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.700%
50) Toluene-d8	10.179	98	250017	47.153	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.300%
62) 4-Bromofluorobenzene	12.483	95	65321	39.200	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.400%
Target Compounds						
17) Carbon Disulfide	4.192	76	5580	1.649	ug/l	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
Data File : VY017670.D
Acq On : 15 Mar 2024 14:12
Operator : SY/MD
Sample : P1746-03
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-4.0-6.0-DUP

Quant Time: Mar 16 00:16:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

