

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
 Data File : VY017676.D
 Acq On : 15 Mar 2024 16:34
 Operator : SY/MD
 Sample : P1746-12
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04-2.0-4.0

Quant Time: Mar 16 00:18:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	71031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	118541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	101340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	35035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	32025	47.573	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.140%
35) Dibromofluoromethane	7.716	113	39915	48.321	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.640%
50) Toluene-d8	10.179	98	134504	46.572	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.140%
62) 4-Bromofluorobenzene	12.483	95	32100	35.366	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	70.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
Data File : VY017676.D
Acq On : 15 Mar 2024 16:34
Operator : SY/MD
Sample : P1746-12
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-04-2.0-4.0

Quant Time: Mar 16 00:18:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

