

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
 Data File : VY017679.D
 Acq On : 15 Mar 2024 17:45
 Operator : SY/MD
 Sample : P1746-02RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-4.0-6.0RE

Quant Time: Mar 16 00:19:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	3731	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	5286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	3691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	476	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1392	39.367	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.740%
35) Dibromofluoromethane	7.728	113	898	24.379	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	48.760%#
50) Toluene-d8	10.185	98	5412	42.023	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.040%
62) 4-Bromofluorobenzene	12.483	95	695	17.171	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	34.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
Data File : VY017679.D
Acq On : 15 Mar 2024 17:45
Operator : SY/MD
Sample : P1746-02RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-4.0-6.0RE

Quant Time: Mar 16 00:19:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

