

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
 Data File : VY017680.D
 Acq On : 15 Mar 2024 18:09
 Operator : SY/MD
 Sample : P1746-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-0-2.0

Quant Time: Mar 16 00:19:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	113143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	207770	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	188967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	70555	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	54068	50.423	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.840%
35) Dibromofluoromethane	7.716	113	68742	47.480	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.960%
50) Toluene-d8	10.179	98	241815	47.770	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.540%
62) 4-Bromofluorobenzene	12.483	95	63706	40.045	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
Data File : VY017680.D
Acq On : 15 Mar 2024 18:09
Operator : SY/MD
Sample : P1746-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-0-2.0

Quant Time: Mar 16 00:19:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

