

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
 Data File : VY017682.D
 Acq On : 15 Mar 2024 18:56
 Operator : SY/MD
 Sample : P1746-06
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-02-0-2.0

Quant Time: Mar 16 00:20:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	94736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	168086	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	157289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	47413	52.808	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.620%	
35) Dibromofluoromethane	7.722	113	57111	48.760	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.520%	
50) Toluene-d8	10.185	98	192835	47.088	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 94.180%	
62) 4-Bromofluorobenzene	12.483	95	55949	43.472	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 86.940%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031524\
Data File : VY017682.D
Acq On : 15 Mar 2024 18:56
Operator : SY/MD
Sample : P1746-06
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-02-0-2.0

Quant Time: Mar 16 00:20:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

