


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031620\
 Data File : VY001970.D
 Acq On : 16 Mar 2020 09:39
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
 Title : SW846 8260
 Last Update : Thu Mar 05 04:05:33 2020



AutoFind: Scans 1764, 1765, 1766; Background Corrected with Scan 1756

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	33037	PASS
75	95	30	60	51.2	86866	PASS
95	95	100	100	100.0	169600	PASS
96	95	5	9	6.8	11488	PASS
173	174	0.00	2	0.8	936	PASS
174	95	50	100	69.6	118117	PASS
175	174	5	9	7.4	8778	PASS
176	174	95	101	96.8	114354	PASS
177	176	5	9	6.9	7871	PASS