

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
 Data File : VY012972.D
 Acq On : 16 Mar 2023 15:40
 Operator : KP/MD
 Sample : 01932-17
 Misc : 8.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JPP-6C-031423

Quant Time: Mar 17 01:36:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	77933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	130646	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	116690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	52450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	50531	65.305	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.600%
35) Dibromofluoromethane	7.716	113	44942	57.001	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.000%
50) Toluene-d8	10.179	98	156985	54.463	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.920%
62) 4-Bromofluorobenzene	12.483	95	47335	48.654	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.300%
Target Compounds						
95) Naphthalene	15.239	128	37583	19.348	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
Data File : VY012972.D
Acq On : 16 Mar 2023 15:40
Operator : KP/MD
Sample : 01932-17
Misc : 8.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JPP-6C-031423

Quant Time: Mar 17 01:36:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

