

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
 Data File : VY012973.D
 Acq On : 16 Mar 2023 16:19
 Operator : KP/MD
 Sample : 01965-01
 Misc : 5.38g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-1

Quant Time: Mar 17 01:37:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	83736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	143064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	131247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62623	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	55719	67.019	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.040%
35) Dibromofluoromethane	7.715	113	48876	56.609	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.220%
50) Toluene-d8	10.178	98	169491	53.698	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.400%
62) 4-Bromofluorobenzene	12.483	95	54789	51.428	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.860%
Target Compounds						
18) Methyl Acetate	4.478	43	7286	7.498	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
Data File : VY012973.D
Acq On : 16 Mar 2023 16:19
Operator : KP/MD
Sample : 01965-01
Misc : 5.38g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-1

Quant Time: Mar 17 01:37:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

