

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
 Data File : VY012976.D
 Acq On : 16 Mar 2023 17:29
 Operator : KP/MD
 Sample : 01964-01
 Misc : 3.82g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 UNION-2

Quant Time: Mar 17 01:38:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	78118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	130197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	118584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	53735	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	50968	65.713	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.420%
35) Dibromofluoromethane	7.716	113	44920	57.169	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.340%
50) Toluene-d8	10.179	98	156380	54.440	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.880%
62) 4-Bromofluorobenzene	12.483	95	47790	49.291	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
Data File : VY012976.D
Acq On : 16 Mar 2023 17:29
Operator : KP/MD
Sample : 01964-01
Misc : 3.82g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
UNION-2

Quant Time: Mar 17 01:38:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

