

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
 Data File : VY012981.D
 Acq On : 16 Mar 2023 19:26
 Operator : KP/MD
 Sample : 01963-01
 Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NYE-SOIL-0316

Quant Time: Mar 17 01:40:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	79320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	133413	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	123788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	53181	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	53357	67.751	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.500%
35) Dibromofluoromethane	7.716	113	46137	57.303	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.600%
50) Toluene-d8	10.179	98	160543	54.542	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.080%
62) 4-Bromofluorobenzene	12.483	95	48270	48.586	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031623\
Data File : VY012981.D
Acq On : 16 Mar 2023 19:26
Operator : KP/MD
Sample : 01963-01
Misc : 5.95g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NYSE-SOIL-0316

Quant Time: Mar 17 01:40:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

