


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031720\
Data File : VY001988.D
Acq On : 17 Mar 2020 09:13
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
Title : SW846 8260
Last Update : Thu Mar 05 04:05:33 2020



AutoFind: Scans 1764, 1765, 1766; Background Corrected with Scan 1756

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	28424	PASS
75	95	30	60	50.3	73168	PASS
95	95	100	100	100.0	145493	PASS
96	95	5	9	6.9	9968	PASS
173	174	0.00	2	0.9	899	PASS
174	95	50	100	72.0	104816	PASS
175	174	5	9	7.5	7903	PASS
176	174	95	101	96.4	101082	PASS
177	176	5	9	7.0	7063	PASS