

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031721\
 Data File : VY004108.D
 Acq On : 17 Mar 2021 17:07
 Operator : SY/MD
 Sample : M1681-39RE
 Misc : 4.52g/5mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 UST3RE

Quant Time: Mar 18 01:51:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	191091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	330887	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	303028	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125000	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	111426	59.09	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.18%
35) Dibromofluoromethane	7.734	113	102748	55.07	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.14%
50) Toluene-d8	10.185	98	391386	53.50	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.00%
62) 4-Bromofluorobenzene	12.483	95	129532	47.94	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.88%
Target Compounds						
69) o-Xylene	12.032	106	7859	2.04	ug/l	92
84) 1,2,4-Trimethylbenzene	13.123	105	7552	1.02	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031721\
Data File : VY004108.D
Acq On : 17 Mar 2021 17:07
Operator : SY/MD
Sample : M1681-39RE
Misc : 4.52g/5mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
UST3RE

Quant Time: Mar 18 01:51:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

