

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031721\
 Data File : VY004105.D
 Acq On : 17 Mar 2021 15:57
 Operator : SY/MD
 Sample : M1691-04
 Misc : 5.94g/5mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FK-GF-02-068-B

Quant Time: Mar 18 01:50:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	194042	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	332906	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	312368	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	144949	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	111705	58.33	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.66%
35) Dibromofluoromethane	7.728	113	104166	55.49	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.98%
50) Toluene-d8	10.185	98	395522	53.74	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.48%
62) 4-Bromofluorobenzene	12.483	95	140106	51.54	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.08%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031721\
Data File : VY004105.D
Acq On : 17 Mar 2021 15:57
Operator : SY/MD
Sample : M1691-04
Misc : 5.94g/5mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FK-GF-02-068-B

Quant Time: Mar 18 01:50:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57




Instrument :
MSVOA_Y
ClientSampleId :
FK-GF-02-068-B


Tgt Ion:168 Resp: 194042
Ion Ratio Lower Upper
168 100
99 53.8 41.2 61.8



#33
1,2-Dichloroethane-d4
Concen: 58.33 ug/l
RT: 8.155 min Scan# 1039
Delta R.T. -0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57

Tgt Ion: 65 Resp: 111705
Ion Ratio Lower Upper
65 100
67 51.7 0.0 101.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57




Instrument :
MSVOA_Y
ClientSampleId :
FK-GF-02-068-B


Tgt Ion:	114	Resp:	332906
Ion	Ratio	Lower	Upper
114	100		
63	19.4	0.0	41.6
88	15.4	0.0	31.4



#35
Dibromofluoromethane
Concen: 55.49 ug/l
RT: 7.728 min Scan# 969
Delta R.T. 0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57

Tgt Ion:	113	Resp:	104166
Ion	Ratio	Lower	Upper
113	100		
111	100.5	83.2	124.8
192	20.4	18.5	27.7









#50
Toluene-d8
Concen: 53.74 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57




Instrument :
MSVOA_Y
ClientSampleId :
FK-GF-02-068-B


Tgt Ion: 98 Resp: 395522
Ion Ratio Lower Upper
98 100
100 66.1 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 51.54 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57

Tgt Ion: 95 Resp: 140106
Ion Ratio Lower Upper
95 100
174 86.4 0.0 179.0
176 83.9 0.0 174.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57

Instrument :
MSVOA_Y
ClientSampleId :
FK-GF-02-068-B

Tgt Ion:117 Resp: 312368
Ion Ratio Lower Upper
117 100
82 53.3 43.8 65.6
119 32.2 26.2 39.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY004105.D
Acq: 17 Mar 2021 15:57

Tgt Ion:152 Resp: 144949
Ion Ratio Lower Upper
152 100
115 57.5 28.2 84.6
150 157.6 0.0 347.2

