

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031723\
 Data File : VY012996.D
 Acq On : 17 Mar 2023 18:56
 Operator : KP/MD
 Sample : 01998-03
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BRAOD-CHANNEL-SUBSTATION-TP2

Quant Time: Mar 18 05:49:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	223	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	468	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	671	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	93	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	410	185.177	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	370.360%#
35) Dibromofluoromethane	7.709	113	115	40.717	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.440%
50) Toluene-d8	10.172	98	775	75.058	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	150.120%#
62) 4-Bromofluorobenzene	12.483	95	100	28.694	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	57.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031723\
Data File : VY012996.D
Acq On : 17 Mar 2023 18:56
Operator : KP/MD
Sample : 01998-03
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BRAOD-CHANNEL-SUBSTATION-TP2

Quant Time: Mar 18 05:49:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

