

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031820\
 Data File : VY002010.D
 Acq On : 18 Mar 2020 17:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 04:31:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 04:05:33 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.462	0.432	6.5	79	0.00
3 P	Chloromethane	0.622	0.613	1.4	82	0.00
4 C	Vinyl Chloride	0.633	0.656	-3.6#	83	0.00
5 T	Bromomethane	0.420	0.437	-4.0	86	0.00
6 T	Chloroethane	0.398	0.432	-8.5	88	0.00
7 T	Trichlorofluoromethane	0.858	0.914	-6.5	87	0.00
8 T	Diethyl Ether	0.338	0.379	-12.1	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.552	0.606	-9.8	90	0.00
10 T	Methyl Iodide	0.638	0.736	-15.4	86	0.00
11 T	Tert butyl alcohol	0.065	0.056	13.8	80	-0.01
12 CM	1,1-Dichloroethene	0.556	0.594	-6.8#	86	0.00
13 T	Acrolein	0.075	0.064	14.7	75	0.00
14 T	Allyl chloride	0.986	1.121	-13.7	91	0.00
15 T	Acrylonitrile	0.181	0.196	-8.3	87	0.00
16 T	Acetone	0.155	0.146	5.8	75	0.00
17 T	Carbon Disulfide	1.869	1.864	0.3	80	0.00
18 T	Methyl Acetate	0.399	0.428	-7.3	86	0.00
19 T	Methyl tert-butyl Ether	1.506	1.693	-12.4	90	0.00
20 T	Methylene Chloride	0.666	0.718	-7.8	92	-0.01
21 T	trans-1,2-Dichloroethene	0.631	0.690	-9.4	88	0.00
22 T	Diisopropyl ether	2.020	2.340	-15.8	93	0.00
23 T	Vinyl Acetate	1.324	1.486	-12.2	89	0.00
24 P	1,1-Dichloroethane	1.094	1.242	-13.5	91	0.00
25 T	2-Butanone	0.236	0.247	-4.7	83	0.00
26 T	2,2-Dichloropropane	0.898	1.025	-14.1	93	0.00
27 T	cis-1,2-Dichloroethene	0.686	0.782	-14.0	91	0.00
28 T	Bromochloromethane	0.493	0.549	-11.4	89	0.00
29 T	Tetrahydrofuran	0.157	0.164	-4.5	84	0.00
30 C	Chloroform	1.028	1.180	-14.8#	92	0.00
31 T	Cyclohexane	1.173	1.204	-2.6	86	0.00
32 T	1,1,1-Trichloroethane	0.866	0.997	-15.1	91	0.00
33 S	1,2-Dichloroethane-d4	0.534	0.562	-5.2	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.273	0.287	-5.1	90	0.00
36 T	1,1-Dichloropropene	0.481	0.518	-7.7	87	0.00
37 T	Ethyl Acetate	0.280	0.296	-5.7	85	0.00
38 T	Carbon Tetrachloride	0.411	0.459	-11.7	91	0.00
39 T	Methylcyclohexane	0.638	0.675	-5.8	87	0.00
40 TM	Benzene	1.426	1.584	-11.1	91	0.00
41 T	Methacrylonitrile	0.126	0.128	-1.6	85	0.00
42 TM	1,2-Dichloroethane	0.367	0.407	-10.9	89	0.00
43 T	Isopropyl Acetate	0.529	0.571	-7.9	87	0.00
44 TM	Trichloroethene	0.347	0.380	-9.5	90	0.00
45 C	1,2-Dichloropropane	0.359	0.405	-12.8#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031820\
 Data File : VY002010.D
 Acq On : 18 Mar 2020 17:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 19 04:31:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 05 04:05:33 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.185	0.202	-9.2	90	0.00
47 T	Bromodichloromethane	0.430	0.496	-15.3	93	0.00
48 T	Methyl methacrylate	0.235	0.264	-12.3	87	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	81	0.00
50 S	Toluene-d8	1.151	1.181	-2.6	87	0.00
51 T	4-Methyl-2-Pentanone	0.273	0.299	-9.5	88	0.00
52 CM	Toluene	0.866	0.963	-11.2#	91	0.00
53 T	t-1,3-Dichloropropene	0.468	0.533	-13.9	90	0.00
54 T	cis-1,3-Dichloropropene	0.557	0.636	-14.2	91	0.00
55 T	1,1,2-Trichloroethane	0.270	0.301	-11.5	90	0.00
56 T	Ethyl methacrylate	0.396	0.437	-10.4	87	0.00
57 T	1,3-Dichloropropane	0.482	0.536	-11.2	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.157	0.134	14.6	81	0.00
59 T	2-Hexanone	0.195	0.198	-1.5	81	0.00
60 T	Dibromochloromethane	0.285	0.330	-15.8	92	0.00
61 T	1,2-Dibromoethane	0.256	0.280	-9.4	89	0.00
62 S	4-Bromofluorobenzene	0.400	0.416	-4.0	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.314	0.343	-9.2	92	0.00
65 PM	Chlorobenzene	0.996	1.122	-12.7	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.327	0.380	-16.2	93	0.00
67 C	Ethyl Benzene	1.858	2.084	-12.2#	91	0.00
68 T	m/p-Xylenes	0.687	0.767	-11.6	90	0.00
69 T	o-Xylene	0.647	0.730	-12.8	91	0.00
70 T	Styrene	1.109	1.268	-14.3	90	0.00
71 P	Bromoform	0.177	0.203	-14.7	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.817	4.316	-13.1	90	0.00
74 T	N-amyl acetate	1.173	1.248	-6.4	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.823	0.897	-9.0	87	0.00
76 T	1,2,3-Trichloropropane	0.593	0.684	-15.3	99	0.00
77 T	Bromobenzene	0.809	0.909	-12.4	91	0.00
78 T	n-propylbenzene	4.709	5.305	-12.7	90	0.00
79 T	2-Chlorotoluene	2.602	2.925	-12.4	90	0.00
80 T	1,3,5-Trimethylbenzene	3.142	3.570	-13.6	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.294	0.324	-10.2	85	0.00
82 T	4-Chlorotoluene	2.709	3.068	-13.3	91	0.00
83 T	tert-Butylbenzene	2.610	2.997	-14.8	92	0.00
84 T	1,2,4-Trimethylbenzene	3.118	3.529	-13.2	90	0.00
85 T	sec-Butylbenzene	3.847	4.403	-14.5	91	0.00
86 T	p-Isopropyltoluene	3.318	3.758	-13.3	90	0.00
87 T	1,3-Dichlorobenzene	1.585	1.772	-11.8	90	0.00
88 T	1,4-Dichlorobenzene	1.608	1.777	-10.5	89	0.00
89 T	n-Butylbenzene	3.443	3.797	-10.3	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv031820\
Data File : VY002010.D
Acq On : 18 Mar 2020 17:19
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 19 04:31:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y030420S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 05 04:05:33 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.626	0.727	-16.1	92	0.00
91 T	1,2-Dichlorobenzene	1.440	1.601	-11.2	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.133	-3.9	85	0.00
93 T	1,2,4-Trichlorobenzene	0.934	0.982	-5.1	85	0.00
94 T	Hexachlorobutadiene	0.455	0.485	-6.6	88	0.00
95 T	Naphthalene	2.215	2.272	-2.6	83	0.00
96 T	1,2,3-Trichlorobenzene	0.822	0.857	-4.3	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6