

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031821\
 Data File : VY004122.D
 Acq On : 18 Mar 2021 15:22
 Operator : SY/MD
 Sample : M1718-01
 Misc : 7.19g/5mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BACKFILL-ASHCROFT-REAR-03

Quant Time: Mar 19 01:43:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	199501	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	346706	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	328039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	162728	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	117259	59.56	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 119.12%		
35) Dibromofluoromethane	7.728	113	106985	54.72	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 109.44%		
50) Toluene-d8	10.185	98	414650	54.10	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 108.20%		
62) 4-Bromofluorobenzene	12.483	95	152342	53.81	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 107.62%		
Target Compounds						
16) Acetone	3.960	43	5566	11.71	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031821\
Data File : VY004122.D
Acq On : 18 Mar 2021 15:22
Operator : SY/MD
Sample : M1718-01
Misc : 7.19g/5mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BACKFILL-ASHCROFT-REAR-03

Quant Time: Mar 19 01:43:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

