

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031821\
 Data File : VY004123.D
 Acq On : 18 Mar 2021 15:47
 Operator : SY/MD
 Sample : M1718-02
 Misc : 5.78g/5mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BACKFILL-ASHCROFT-REAR-04

Quant Time: Mar 19 01:43:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	193572	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	334780	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	323518	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	162457	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	112501	58.89	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.78%
35) Dibromofluoromethane	7.728	113	104091	55.14	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.28%
50) Toluene-d8	10.185	98	403120	54.47	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.94%
62) 4-Bromofluorobenzene	12.483	95	148745	54.42	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.84%

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031821\
Data File : VY004123.D
Acq On : 18 Mar 2021 15:47
Operator : SY/MD
Sample : M1718-02
Misc : 5.78g/5mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BACKFILL-ASHCROFT-REAR-04

Quant Time: Mar 19 01:43:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

