

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031821\
 Data File : VY004113.D
 Acq On : 18 Mar 2021 11:11
 Operator : SY/MD
 Sample : VY0318SBL01
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0318SBL01

Quant Time: Mar 19 01:41:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	202072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	349084	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	329666	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	161540	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	110779	55.55	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.10%
35) Dibromofluoromethane	7.728	113	106939	54.32	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.64%
50) Toluene-d8	10.185	98	414147	53.66	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.32%
62) 4-Bromofluorobenzene	12.483	95	149485	52.45	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.90%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031821\
Data File : VY004113.D
Acq On : 18 Mar 2021 11:11
Operator : SY/MD
Sample : VY0318SBL01
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0318SBL01

Quant Time: Mar 19 01:41:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.000 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11




Instrument :
MSVOA_Y
ClientSampleId :
VY0318SBL01


Tgt Ion:168 Resp: 202072
Ion Ratio Lower Upper
168 100
99 52.7 41.2 61.8



#33
1,2-Dichloroethane-d4
Concen: 55.55 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. -0.006 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11

Tgt Ion: 65 Resp: 110779
Ion Ratio Lower Upper
65 100
67 51.0 0.0 101.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11

Instrument :
MSVOA_Y
ClientSampleId :
VY0318SBL01

Tgt Ion:114 Resp: 349084




Ion	Ratio	Lower	Upper
114	100		
63	18.0	0.0	41.6
88	14.2	0.0	31.4




#35
Dibromofluoromethane
Concen: 54.32 ug/l
RT: 7.728 min Scan# 969
Delta R.T. -0.000 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11

Tgt Ion:113 Resp: 106939

Ion	Ratio	Lower	Upper
113	100		
111	103.8	83.2	124.8
192	20.2	18.5	27.7









#50
Toluene-d8
Concen: 53.66 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11




Instrument :
MSVOA_Y
ClientSampleId :
VY0318SBL01


Tgt Ion: 98 Resp: 414147
Ion Ratio Lower Upper
98 100
100 66.0 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 52.45 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11

Tgt Ion: 95 Resp: 149485
Ion Ratio Lower Upper
95 100
174 86.0 0.0 179.0
176 83.6 0.0 174.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11

Instrument :
MSVOA_Y
ClientSampleId :
VY0318SBL01

Tgt Ion:	117	Resp:	329666
Ion	Ratio	Lower	Upper
117	100		
82	51.8	43.8	65.6
119	32.2	26.2	39.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY004113.D
Acq: 18 Mar 2021 11:11

Tgt Ion:	152	Resp:	161540
Ion	Ratio	Lower	Upper
152	100		
115	57.2	28.2	84.6
150	159.6	0.0	347.2

