



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031821\
 Data File : VY004111.D
 Acq On : 18 Mar 2021 09:45
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Title : SW846 8260
 Last Update : Mon Mar 08 12:28:55 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	14700	PASS
75	95	30	60	52.7	41048	PASS
95	95	100	100	100.0	77896	PASS
96	95	5	9	6.9	5367	PASS
173	174	0.00	2	1.1	716	PASS
174	95	50	100	83.9	65336	PASS
175	174	5	9	7.7	5055	PASS
176	174	95	101	96.5	63072	PASS
177	176	5	9	6.5	4085	PASS