

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031825\
 Data File : VY021557.D
 Acq On : 18 Mar 2025 11:06
 Operator : SY/MD
 Sample : VY0318SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0318SBL01

Quant Time: Mar 19 01:38:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	252095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	463980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	410354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	156201	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.060	65	148257	55.642	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.280%
35) Dibromofluoromethane	7.634	113	156514	51.319	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.640%
50) Toluene-d8	10.109	98	567845	49.175	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.407	95	166006	42.262	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031825\
Data File : VY021557.D
Acq On : 18 Mar 2025 11:06
Operator : SY/MD
Sample : VY0318SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0318SBL01

Quant Time: Mar 19 01:38:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

