

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
 Data File : VY004137.D
 Acq On : 19 Mar 2021 14:25
 Operator : SY/MD
 Sample : M1723-01RE
 Misc : 5.56g/5mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 24055RE

Quant Time: Mar 20 03:14:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	78571	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	104541	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	61688	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	22104	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	18323	23.63	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	47.26%#
35) Dibromofluoromethane	7.728	113	24417	41.42	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.84%
50) Toluene-d8	10.185	98	113317	49.03	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.06%
62) 4-Bromofluorobenzene	12.483	95	22799	26.71	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	53.42%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
Data File : VY004137.D
Acq On : 19 Mar 2021 14:25
Operator : SY/MD
Sample : M1723-01RE
Misc : 5.56g/5mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
24055RE

Quant Time: Mar 20 03:14:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

