

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
 Data File : VY004138.D
 Acq On : 19 Mar 2021 15:35
 Operator : SY/MD
 Sample : M1721-01
 Misc : 3.34g/5mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-031821

Quant Time: Mar 20 03:14:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	8599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	15085	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	12820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4708	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	5251	61.88	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.76%
35) Dibromofluoromethane	7.728	113	2889	33.96	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	67.92%
50) Toluene-d8	10.185	98	17046	51.11	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.22%
62) 4-Bromofluorobenzene	12.483	95	5553	45.08	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.16%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.117	105	880	3.17	ug/l	Qvalue 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
Data File : VY004138.D
Acq On : 19 Mar 2021 15:35
Operator : SY/MD
Sample : M1721-01
Misc : 3.34g/5mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-031821

Quant Time: Mar 20 03:14:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

