

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
 Data File : VY004139.D
 Acq On : 19 Mar 2021 15:58
 Operator : SY/MD
 Sample : M1721-03
 Misc : 3.35g/5mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-08-031821

Quant Time: Mar 20 03:14:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	53211	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	78704	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	59923	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	23954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	20802	39.61	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	79.22%		
35) Dibromofluoromethane	7.728	113	20634	46.49	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	92.98%		
50) Toluene-d8	10.185	98	88588	50.91	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	101.82%		
62) 4-Bromofluorobenzene	12.483	95	24741	38.50	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	77.00%		
Target Compounds						Qvalue
16) Acetone	3.961	43	1650	13.02	ug/l	99
67) Ethyl Benzene	11.599	91	4641	2.20	ug/l	92
68) m/p-Xylenes	11.703	106	9818	12.21	ug/l	92
69) o-Xylene	12.032	106	3548	4.67	ug/l	90
84) 1,2,4-Trimethylbenzene	13.123	105	2340	1.65	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
Data File : VY004139.D
Acq On : 19 Mar 2021 15:58
Operator : SY/MD
Sample : M1721-03
Misc : 3.35g/5mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-08-031821

Quant Time: Mar 20 03:14:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

