

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
 Data File : VY004145.D
 Acq On : 19 Mar 2021 19:04
 Operator : SY/MD
 Sample : M1746-04
 Misc : 5.87g/5mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-4-18.5-19.0

Quant Time: Mar 20 03:16:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	150982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	248253	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	227906	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	111638	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	80158	53.80	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.60%	
35) Dibromofluoromethane	7.734	113	74429	53.17	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.34%	
50) Toluene-d8	10.185	98	291288	53.07	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.14%	
62) 4-Bromofluorobenzene	12.483	95	104141	51.38	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.76%	
Target Compounds						
16) Acetone	3.966	43	5173	14.38	ug/l	91
20) Methylene Chloride	4.735	84	11229	3.05	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031921\
Data File : VY004145.D
Acq On : 19 Mar 2021 19:04
Operator : SY/MD
Sample : M1746-04
Misc : 5.87g/5mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-4-18.5-19.0

Quant Time: Mar 20 03:16:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

