

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031924\
 Data File : VY017703.D
 Acq On : 19 Mar 2024 11:16
 Operator : SY/MD
 Sample : VY0319SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0319SBL01

Quant Time: Mar 19 11:59:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	141623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.704	114	234978	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.502	117	203836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	84592	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	63644	47.418	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.840%
35) Dibromofluoromethane	7.734	113	73452	44.859	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.720%
50) Toluene-d8	10.191	98	265651	46.402	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.800%
62) 4-Bromofluorobenzene	12.496	95	71452	39.713	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031924\
Data File : VY017703.D
Acq On : 19 Mar 2024 11:16
Operator : SY/MD
Sample : VY0319SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0319SBL01

Quant Time: Mar 19 11:59:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

