

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031924\
 Data File : VY017704.D
 Acq On : 19 Mar 2024 11:39
 Operator : SY/MD
 Sample : VY0319SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0319SBL02

Quant Time: Mar 19 12:00:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	132462	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.709	114	221088	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.508	117	193418	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.440	152	77968	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	60064	47.845	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.700%
35) Dibromofluoromethane	7.734	113	69686	45.233	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.460%
50) Toluene-d8	10.197	98	250873	46.574	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.140%
62) 4-Bromofluorobenzene	12.495	95	67354	39.787	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY031924\
Data File : VY017704.D
Acq On : 19 Mar 2024 11:39
Operator : SY/MD
Sample : VY0319SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0319SBL02

Quant Time: Mar 19 12:00:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

