

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032023\
 Data File : VY013015.D
 Acq On : 20 Mar 2023 18:01
 Operator : KP/MD
 Sample : 02002-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 031023-3

Quant Time: Mar 21 01:15:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	75485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	128016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	123899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	56539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	52590	70.170	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.340%	
35) Dibromofluoromethane	7.716	113	46354	59.999	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 120.000%	
50) Toluene-d8	10.179	98	157185	55.653	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.300%	
62) 4-Bromofluorobenzene	12.483	95	49058	51.461	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.920%	
Target Compounds						
25) 2-Butanone	6.990	43	3680	10.100	ug/l	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032023\
Data File : VY013015.D
Acq On : 20 Mar 2023 18:01
Operator : KP/MD
Sample : 02002-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
031023-3

Quant Time: Mar 21 01:15:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

