

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032024\
 Data File : VY017724.D
 Acq On : 20 Mar 2024 12:02
 Operator : SY/MD
 Sample : P1804-03
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JC-03-03192024-B

Quant Time: Mar 21 03:14:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	109244	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	209690	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.496	117	185235	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	67639	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	58460	56.465	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.920%
35) Dibromofluoromethane	7.734	113	68028	46.557	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.120%
50) Toluene-d8	10.191	98	245435	48.041	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.080%
62) 4-Bromofluorobenzene	12.489	95	64031	39.880	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032024\
Data File : VY017724.D
Acq On : 20 Mar 2024 12:02
Operator : SY/MD
Sample : P1804-03
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JC-03-03192024-B

Quant Time: Mar 21 03:14:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

