

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032024\
 Data File : VY017722.D
 Acq On : 20 Mar 2024 10:40
 Operator : SY/MD
 Sample : VY0320SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024

Quant Time: Mar 21 03:12:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	134807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	218597	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.502	117	186450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	84144	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	53452	41.838	ug/l	0.01
Spiked Amount 50.000	Range 50 - 163		Recovery =	83.680%		
35) Dibromofluoromethane	7.728	113	63883	41.939	ug/l	0.01
Spiked Amount 50.000	Range 54 - 147		Recovery =	83.880%		
50) Toluene-d8	10.191	98	241884	45.417	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	90.840%		
62) 4-Bromofluorobenzene	12.495	95	72577	43.361	ug/l	0.01
Spiked Amount 50.000	Range 30 - 143		Recovery =	86.720%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	18168	18.270	ug/l	98
3) Chloromethane	2.119	50	33044	18.187	ug/l	100
4) Vinyl Chloride	2.253	62	45656	19.587	ug/l	99
5) Bromomethane	2.656	94	37865	19.934	ug/l	97
6) Chloroethane	2.802	64	30665	20.854	ug/l	98
7) Trichlorofluoromethane	3.137	101	44326	18.658	ug/l	96
8) Diethyl Ether	3.533	74	12209	17.522	ug/l	93
9) 1,1,2-Trichlorotrifluo...	3.905	101	28525	20.875	ug/l	98
10) Methyl Iodide	4.100	142	29059	18.949	ug/l	99
11) Tert butyl alcohol	4.960	59	8657	79.770	ug/l	98
12) 1,1-Dichloroethene	3.881	96	25946	20.374	ug/l	99
13) Acrolein	3.735	56	9341	65.233	ug/l	94
14) Allyl chloride	4.484	41	31179	20.263	ug/l	99
15) Acrylonitrile	5.173	53	25538	89.159	ug/l	95
16) Acetone	3.954	43	23860	96.286	ug/l #	85
17) Carbon Disulfide	4.210	76	72191	18.886	ug/l	98
18) Methyl Acetate	4.491	43	9366	15.579	ug/l	98
19) Methyl tert-butyl Ether	5.228	73	53087	17.885	ug/l	96
20) Methylene Chloride	4.728	84	31501	18.147	ug/l	99
21) trans-1,2-Dichloroethene	5.228	96	29445	19.748	ug/l	96
22) Diisopropyl ether	6.131	45	72734	20.484	ug/l	98
23) Vinyl Acetate	6.070	43	198429	93.782	ug/l	97
24) 1,1-Dichloroethane	6.033	63	50003	20.613	ug/l	96
25) 2-Butanone	6.996	43	32276	91.394	ug/l	92
26) 2,2-Dichloropropane	6.996	77	43137	21.150	ug/l	98
27) cis-1,2-Dichloroethene	7.002	96	32561	19.450	ug/l	96
28) Bromochloromethane	7.344	49	17062	17.859	ug/l	98
29) Tetrahydrofuran	7.362	42	17461	82.022	ug/l	97
30) Chloroform	7.514	83	52863	20.468	ug/l	100
31) Cyclohexane	7.795	56	43564	20.952	ug/l	98
32) 1,1,1-Trichloroethane	7.716	97	46083	20.552	ug/l	97
36) 1,1-Dichloropropene	7.929	75	41014	19.956	ug/l	99
37) Ethyl Acetate	7.082	43	13483	16.610	ug/l	98
38) Carbon Tetrachloride	7.911	117	39968	19.404	ug/l	99
39) Methylcyclohexane	9.197	83	47982	19.427	ug/l	95
40) Benzene	8.173	78	119058	19.507	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032024\
 Data File : VY017722.D
 Acq On : 20 Mar 2024 10:40
 Operator : SY/MD
 Sample : VY0320SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024

Quant Time: Mar 21 03:12:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.319	41	6142m	15.078	ug/l	
42) 1,2-Dichloroethane	8.246	62	27436	18.424	ug/l	100
43) Isopropyl Acetate	8.283	43	23864	16.730	ug/l	96
44) Trichloroethene	8.947	130	33242	18.896	ug/l	98
45) 1,2-Dichloropropane	9.228	63	28352	19.525	ug/l	98
46) Dibromomethane	9.319	93	14687	17.673	ug/l	98
47) Bromodichloromethane	9.508	83	38669	19.164	ug/l	99
48) Methyl methacrylate	9.301	41	11098	16.809	ug/l	99
49) 1,4-Dioxane	9.307	88	2780	295.589	ug/l #	98
51) 4-Methyl-2-Pentanone	10.081	43	62632	78.784	ug/l	100
52) Toluene	10.258	92	74918	19.471	ug/l	100
53) t-1,3-Dichloropropene	10.477	75	33650	17.967	ug/l	97
54) cis-1,3-Dichloropropene	9.941	75	42521	18.685	ug/l	98
55) 1,1,2-Trichloroethane	10.654	97	20624	17.479	ug/l	95
56) Ethyl methacrylate	10.520	69	21384	16.521	ug/l	96
57) 1,3-Dichloropropane	10.800	76	34955	18.221	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.794	63	59864	87.923	ug/l	100
59) 2-Hexanone	10.843	43	44873	82.176	ug/l	99
60) Dibromochloromethane	10.996	129	24892	17.833	ug/l	99
61) 1,2-Dibromoethane	11.099	107	18614	17.119	ug/l	99
64) Tetrachloroethene	10.733	164	33093	18.492	ug/l	98
65) Chlorobenzene	11.526	112	76451	19.318	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.599	131	28690	19.351	ug/l	98
67) Ethyl Benzene	11.605	91	138841	20.004	ug/l	97
68) m/p-Xylenes	11.715	106	107560	39.439	ug/l	99
69) o-Xylene	12.044	106	49553	19.585	ug/l	97
70) Styrene	12.056	104	79885	19.021	ug/l	99
71) Bromoform	12.221	173	13969	17.002	ug/l #	98
73) Isopropylbenzene	12.343	105	130172	20.823	ug/l	99
74) N-amyl acetate	12.154	43	19382	17.579	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.593	83	20716	19.298	ug/l	99
76) 1,2,3-Trichloropropane	12.642	75	15298m	19.135	ug/l	
77) Bromobenzene	12.623	156	29500	19.486	ug/l	97
78) n-propylbenzene	12.684	91	161375	21.514	ug/l	100
79) 2-Chlorotoluene	12.770	91	87406	20.813	ug/l	100
80) 1,3,5-Trimethylbenzene	12.824	105	104759	20.917	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.392	75	6257	17.983	ug/l	97
82) 4-Chlorotoluene	12.867	91	89896	20.928	ug/l	100
83) tert-Butylbenzene	13.087	119	91425	20.788	ug/l	99
84) 1,2,4-Trimethylbenzene	13.135	105	101440	20.516	ug/l	99
85) sec-Butylbenzene	13.263	105	143537	21.347	ug/l	100
86) p-Isopropyltoluene	13.379	119	115925	21.061	ug/l	99
87) 1,3-Dichlorobenzene	13.379	146	59622	19.621	ug/l	100
88) 1,4-Dichlorobenzene	13.458	146	58617	19.842	ug/l	99
89) n-Butylbenzene	13.708	91	106387	21.185	ug/l	98
90) Hexachloroethane	13.971	117	23169	20.996	ug/l	98
91) 1,2-Dichlorobenzene	13.751	146	49368	19.207	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.367	75	2478	16.420	ug/l	93
93) 1,2,4-Trichlorobenzene	15.019	180	25404	18.712	ug/l	97
94) Hexachlorobutadiene	15.123	225	17219	20.542	ug/l	98
95) Naphthalene	15.251	128	37119	16.920	ug/l	100
96) 1,2,3-Trichlorobenzene	15.440	180	21282	18.273	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032024\
Data File : VY017722.D
Acq On : 20 Mar 2024 10:40
Operator : SY/MD
Sample : VY0320SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024

Quant Time: Mar 21 03:12:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Quant Title : SW846 8260
QLast Update : Sat Feb 24 04:41:20 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032024\
 Data File : VY017722.D
 Acq On : 20 Mar 2024 10:40
 Operator : SY/MD
 Sample : VY0320SBSD01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBSD01

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024


Quant Time: Mar 21 03:12:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 24 04:41:20 2024
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.801 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01



Tgt Ion:168 Resp: 13480
Ion Ratio Lower Upper
168 100
99 51.9 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#2
Dichlorodifluoromethane
Concen: 18.270 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40


Tgt Ion: 85 Resp: 18168
Ion Ratio Lower Upper
85 100
87 31.1 16.0 48.0





#3
Chloromethane
Concen: 18.187 ug/l
RT: 2.119 min Scan# 49
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01




Tgt Ion: 50 Resp: 3304
Ion Ratio Lower Upper
50 100
52 31.9 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024




#4
Vinyl Chloride
Concen: 19.587 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40



Tgt Ion: 62 Resp: 45656
Ion Ratio Lower Upper
62 100
64 32.0 26.1 39.1









#5
Bromomethane
Concen: 19.934 ug/l
RT: 2.656 min Scan# 111
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion: 94 Resp: 3786
Ion Ratio Lower Upper
94 100
96 89.6 74.1 111.1




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#6
Chloroethane
Concen: 20.854 ug/l
RT: 2.802 min Scan# 161
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 64 Resp: 30665
Ion Ratio Lower Upper
64 100
66 31.6 26.1 39.1









#7
 Trichlorofluoromethane
 Concen: 18.658 ug/l
 RT: 3.137 min Scan# 213
 Delta R.T. 0.018 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBS01

Tgt Ion:101 Resp: 44320
 Ion Ratio Lower Upper
 101 100
 103 65.0 49.3 73.9




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024



#8
 Diethyl Ether
 Concen: 17.522 ug/l
 RT: 3.533 min Scan# 281
 Delta R.T. 0.006 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40

Tgt Ion: 74 Resp: 12209
 Ion Ratio Lower Upper
 74 100
 45 88.6 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.875 ug/l
RT: 3.905 min Scan# 340
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion:101 Resp: 2852
Ion Ratio Lower Upper
101 100
85 43.8 34.5 51.7
151 74.7 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#10
Methyl Iodide
Concen: 18.949 ug/l
RT: 4.100 min Scan# 374
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:142 Resp: 29059
Ion Ratio Lower Upper
142 100
127 40.5 32.8 49.2
141 14.5 11.4 17.2









#11
Tert butyl alcohol
Concen: 79.770 ug/l
RT: 4.960 min Scan# 514
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion: 59 Resp: 865
Ion Ratio Lower Upper
59 100
57 4.3 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#12
1,1-Dichloroethene
Concen: 20.374 ug/l
RT: 3.881 min Scan# 338
Delta R.T. 0.018 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 96 Resp: 25946
Ion Ratio Lower Upper
96 100
61 144.0 114.6 171.8
98 65.7 50.8 76.2





#13

Acrolein

Concen: 65.233 ug/l

RT: 3.735 min Scan# 312

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VY0320SBS01

Tgt Ion: 56 Resp: 934

Ion Ratio Lower Upper

56 100





55 64.6 55.7 83.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024

Supervised By :Semsettin Yesilyurt 03/21/2024



#14

Allyl chloride

Concen: 20.263 ug/l

RT: 4.484 min Scan# 437

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40




Tgt Ion: 41 Resp: 31179


Ion Ratio Lower Upper

41 100

39 73.0 58.3 87.5

76 48.0 39.0 58.6









#15
Acrylonitrile
Concen: 89.159 ug/l
RT: 5.173 min Scan# 51
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01

Tgt Ion: 53 Resp: 25538
Ion Ratio Lower Upper
53 100
52 75.3 64.7 97.1
51 34.5 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#16
Acetone
Concen: 96.286 ug/l
RT: 3.954 min Scan# 350
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 43 Resp: 23860
Ion Ratio Lower Upper
43 100
58 30.2 31.7 47.5#









#17
Carbon Disulfide
Concen: 18.886 ug/l
RT: 4.210 min Scan# 392
Delta R.T. 0.018 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01

Tgt Ion: 76 Resp: 7219
Ion Ratio Lower Upper
76 100
78 9.1 6.8 10.2




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#18
Methyl Acetate
Concen: 15.579 ug/l
RT: 4.491 min Scan# 438
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 43 Resp: 9366
Ion Ratio Lower Upper
43 100
74 28.3 23.3 34.9









#19
Methyl tert-butyl Ether
Concen: 17.885 ug/l
RT: 5.228 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01

Tgt Ion: 73 Resp: 5308
Ion Ratio Lower Upper
73 100
57 19.9 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#20
Methylene Chloride
Concen: 18.147 ug/l
RT: 4.728 min Scan# 477
Delta R.T. 0.018 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40


Tgt Ion: 84 Resp: 31501
Ion Ratio Lower Upper
84 100
49 105.7 84.4 126.6
51 32.5 26.0 39.0
86 63.0 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 19.748 ug/l
RT: 5.228 min Scan# 511
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01




Tgt Ion: 96 Resp: 2944
Ion Ratio Lower Upper
96 100
61 125.6 97.5 146.3
98 66.2 50.5 75.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024




#22
Diisopropyl ether
Concen: 20.484 ug/l
RT: 6.131 min Scan# 707
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40



Tgt Ion: 45 Resp: 72734
Ion Ratio Lower Upper
45 100
43 54.2 42.2 63.2
87 32.1 25.0 37.6
59 10.8 10.5 15.7









#23
 Vinyl Acetate
 Concen: 93.782 ug/l
 RT: 6.070 min Scan# 695
 Delta R.T. 0.012 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBSD01

Tgt Ion: 43 Resp: 198429
 Ion Ratio Lower Upper
 43 100
 86 12.3 10.6 16.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024



#24
 1,1-Dichloroethane
 Concen: 20.613 ug/l
 RT: 6.033 min Scan# 691
 Delta R.T. 0.018 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40


Tgt Ion: 63 Resp: 50003
 Ion Ratio Lower Upper
 63 100
 98 6.3 4.0 12.2
 100 4.1 2.6 8.0





#25
2-Butanone
Concen: 91.394 ug/l
RT: 6.996 min Scan# 849
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01




Tgt Ion: 43 Resp: 32270
Ion Ratio Lower Upper
43 100
72 29.6 27.3 40.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024




#26
2,2-Dichloropropane
Concen: 21.150 ug/l
RT: 6.996 min Scan# 849
Delta R.T. 0.018 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Tgt Ion: 77 Resp: 43137
Ion Ratio Lower Upper
77 100
97 23.2 12.0 36.1





#27
 cis-1,2-Dichloroethene
 Concen: 19.450 ug/l
 RT: 7.002 min Scan# 848
 Delta R.T. 0.012 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBS01



Tgt Ion: 96 Resp: 3256
 Ion Ratio Lower Upper
 96 100
 61 134.9 0.0 256.2
 98 64.4 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024



#28
 Bromochloromethane
 Concen: 17.859 ug/l
 RT: 7.344 min Scan# 906
 Delta R.T. 0.006 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40

Tgt Ion: 49 Resp: 17062
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 4.6
 130 89.2 72.7 109.1





#29

Tetrahydrofuran

Concen: 82.022 ug/l

RT: 7.362 min Scan# 907

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion: 42 Resp: 17463
 Ion Ratio Lower Upper
 42 100
 72 51.8 43.0 64.4
 71 48.0 40.8 61.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024



#30

Chloroform

Concen: 20.468 ug/l




RT: 7.514 min Scan# 934


Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion: 83 Resp: 52863
 Ion Ratio Lower Upper
 83 100
 85 64.9 52.2 78.4









#31
Cyclohexane
Concen: 20.952 ug/l
RT: 7.795 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion: 56 Resp: 43564
Ion Ratio Lower Upper
56 100
69 36.5 28.7 43.1
84 100.0 77.7 116.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#32
1,1,1-Trichloroethane
Concen: 20.552 ug/l
RT: 7.716 min Scan# 967
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 97 Resp: 46083
Ion Ratio Lower Upper
97 100
99 64.7 50.3 75.5
61 42.6 32.6 48.8





#33

1,2-Dichloroethane-d4

Concen: 41.838 ug/l

RT: 8.155 min Scan# 1109

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VY0320SBS01

Tgt Ion: 65 Resp: 53452

Ion Ratio Lower Upper

65 100





67 56.3 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024

Supervised By :Semsettin Yesilyurt 03/21/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.703 min Scan# 1129

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40




Tgt Ion:114 Resp: 218597

Ion Ratio Lower Upper

114 100

63 17.3 0.0 35.0

88 14.0 0.0 27.2









#35
Dibromofluoromethane
Concen: 41.939 ug/l
RT: 7.728 min Scan# 90
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01

Tgt Ion:113 Resp: 6388
Ion Ratio Lower Upper
113 100
111 103.1 82.2 123.4
192 20.3 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#36
1,1-Dichloropropene
Concen: 19.956 ug/l
RT: 7.929 min Scan# 1002
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 75 Resp: 41014
Ion Ratio Lower Upper
75 100
110 37.3 19.0 57.0
77 31.1 25.3 37.9





#37

Ethyl Acetate

Concen: 16.610 ug/l

RT: 7.082 min Scan# 80

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VY0320SBS01





Tgt Ion: 43 Resp: 1348

Ion	Ratio	Lower	Upper
43	100		
61	15.5	12.6	19.0
70	11.1	7.8	11.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#38

Carbon Tetrachloride

Concen: 19.404 ug/l

RT: 7.911 min Scan# 999




Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion:117 Resp: 39968

Ion	Ratio	Lower	Upper
117	100		
119	96.3	75.8	113.6
121	29.6	24.0	36.0





#39

Methylcyclohexane

Concen: 19.427 ug/l

RT: 9.197 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VY0320SBS01

Tgt Ion: 83 Resp: 4798

Ion Ratio Lower Upper

83 100

55 68.6 51.5 77.3





98 47.3 40.5 60.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024

Supervised By :Semsettin Yesilyurt 03/21/2024



#40

Benzene

Concen: 19.507 ug/l

RT: 8.173 min Scan# 1042

Delta R.T. 0.012 min

Lab File: VY017722.D




Acq: 20 Mar 2024 10:40


Tgt Ion: 78 Resp: 119058

Ion Ratio Lower Upper

78 100


77 23.0 19.5 29.3





#41
Methacrylonitrile
Concen: 15.078 ug/l m
RT: 7.319 min Scan# 903
Delta R.T. -0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01



Tgt Ion: 41 Resp: 6142
Ion Ratio Lower Upper
41 100
39 0.0 48.0 72.0
67 0.0 75.1 112.7
52 0.0 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#42
1,2-Dichloroethane
Concen: 18.424 ug/l
RT: 8.246 min Scan# 1054
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40


Tgt Ion: 62 Resp: 27436
Ion Ratio Lower Upper
62 100
98 10.8 0.0 22.0





#43
Isopropyl Acetate
Concen: 16.730 ug/l
RT: 8.283 min Scan# 1060
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01




Tgt Ion: 43 Resp: 2386
Ion Ratio Lower Upper
43 100
61 37.6 27.9 41.9
87 15.8 13.1 19.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024




#44
Trichloroethene
Concen: 18.896 ug/l
RT: 8.947 min Scan# 1169
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40



Tgt Ion:130 Resp: 33242
Ion Ratio Lower Upper
130 100
95 92.2 0.0 188.0





#45

1,2-Dichloropropane

Concen: 19.525 ug/l

RT: 9.228 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VY0320SBS01

Tgt Ion: 63 Resp: 2835

Ion Ratio Lower Upper

63 100





65 32.9 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024

Supervised By :Semsettin Yesilyurt 03/21/2024



#46

Dibromomethane

Concen: 17.673 ug/l

RT: 9.319 min Scan# 1230

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40




Tgt Ion: 93 Resp: 14687


Ion Ratio Lower Upper

93 100

95 83.4 66.2 99.4

174 104.6 80.9 121.3





#47

Bromodichloromethane

Concen: 19.164 ug/l

RT: 9.508 min Scan# 1261

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion:	83	Resp:	3866
Ion Ratio	Lower	Upper	
83	100		
85	63.4	50.1	75.1
127	8.4	7.4	11.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#48

Methyl methacrylate

Concen: 16.809 ug/l




RT: 9.301 min Scan# 1227


Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion:	41	Resp:	11098
Ion Ratio	Lower	Upper	
41	100		
69	105.6	84.1	126.1
39	55.1	43.6	65.4





#49

1,4-Dioxane

Concen: 295.589 ug/l

RT: 9.307 min Scan# 1227

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VY0320SBS01

Tgt Ion: 88 Resp: 2780

Ion Ratio Lower Upper

88 100

43 29.5 0.0 0.0





58 59.7 49.3 73.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024

Supervised By :Semsettin Yesilyurt 03/21/2024



#50

Toluene-d8

Concen: 45.417 ug/l

RT: 10.191 min Scan# 1373

Delta R.T. 0.006 min

Lab File: VY017722.D




Acq: 20 Mar 2024 10:40


Tgt Ion: 98 Resp: 241884

Ion Ratio Lower Upper

98 100


100 64.1 52.0 78.0





#51
4-Methyl-2-Pentanone
Concen: 78.784 ug/l
RT: 10.081 min Scan# 1355
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01




Tgt Ion: 43 Resp: 62632
Ion Ratio Lower Upper
43 100
58 44.6 35.7 53.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024




#52
Toluene
Concen: 19.471 ug/l
RT: 10.258 min Scan# 1384
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Tgt Ion: 92 Resp: 74918
Ion Ratio Lower Upper
92 100
91 174.3 139.0 208.6





#53
 t-1,3-Dichloropropene
 Concen: 17.967 ug/l
 RT: 10.477 min Scan# 1419
 Delta R.T. 0.006 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0320SBS01




Tgt Ion: 75 Resp: 33650
 Ion Ratio Lower Upper
 75 100
 77 31.6 24.1 36.1

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024




#54
 cis-1,3-Dichloropropene
 Concen: 18.685 ug/l
 RT: 9.941 min Scan# 1332
 Delta R.T. 0.006 min
 Lab File: VY017722.D
 Acq: 20 Mar 2024 10:40




Tgt Ion: 75 Resp: 42521
 Ion Ratio Lower Upper
 75 100
 77 32.7 25.4 38.0
 39 36.4 29.8 44.6





#55
1,1,2-Trichloroethane
Concen: 17.479 ug/l
RT: 10.654 min Scan# 1448
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01



Tgt Ion	Ratio	Lower	Upper
97	100		
83	85.3	67.5	101.3
85	55.9	42.9	64.3
99	66.8	46.6	70.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#56
Ethyl methacrylate
Concen: 16.521 ug/l
RT: 10.520 min Scan# 1427
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion	Ratio	Lower	Upper
69	100		
41	55.8	42.3	63.5
39	27.2	22.9	34.3





#57

1,3-Dichloropropane

Concen: 18.221 ug/l

RT: 10.800 min Scan# 1472

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion: 76 Resp: 3495
 Ion Ratio Lower Upper
 76 100
 78 32.4 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024



#58

2-Chloroethyl Vinyl ether

Concen: 87.923 ug/l




RT: 9.794 min Scan# 1308


Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40


Tgt Ion: 63 Resp: 59864
 Ion Ratio Lower Upper
 63 100
 106 29.4 23.7 35.5





#59
2-Hexanone
Concen: 82.176 ug/l
RT: 10.843 min Scan# 1479
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01



Tgt Ion: 43 Resp: 4487
Ion Ratio Lower Upper
43 100
58 62.3 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#60
Dibromochloromethane
Concen: 17.833 ug/l
RT: 10.996 min Scan# 1505
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:129 Resp: 24892
Ion Ratio Lower Upper
129 100
127 78.7 39.0 117.0









#61
1,2-Dibromoethane
Concen: 17.119 ug/l
RT: 11.099 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01

Tgt Ion:107 Resp: 18614
Ion Ratio Lower Upper
107 100
109 96.2 75.9 113.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#62
4-Bromofluorobenzene
Concen: 43.361 ug/l
RT: 12.495 min Scan# 1751
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40


Tgt Ion: 95 Resp: 72577
Ion Ratio Lower Upper
95 100
174 89.8 0.0 175.6
176 85.8 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.502 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01



Tgt Ion:117 Resp: 186450
Ion Ratio Lower Upper
117 100
82 52.1 42.4 63.6
119 32.3 25.9 38.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#64
Tetrachloroethene
Concen: 18.492 ug/l
RT: 10.733 min Scan# 1462
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:164 Resp: 33093
Ion Ratio Lower Upper
164 100
166 132.6 102.6 154.0
129 90.4 72.0 108.0
131 89.7 70.0 105.0









#65
Chlorobenzene
Concen: 19.318 ug/l
RT: 11.526 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion:112 Resp: 7645
Ion Ratio Lower Upper
112 100
114 32.7 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 19.351 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40


Tgt Ion:131 Resp: 28690
Ion Ratio Lower Upper
131 100
133 93.8 47.8 143.3
119 66.0 34.1 102.2





#67
Ethyl Benzene
Concen: 20.004 ug/l
RT: 11.605 min Scan# 1605
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01



Tgt Ion: 91 Resp: 13884
Ion Ratio Lower Upper
91 100
106 32.3 24.5 36.7




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#68
m/p-Xylenes
Concen: 39.439 ug/l
RT: 11.715 min Scan# 1623
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:106 Resp: 107560
Ion Ratio Lower Upper
106 100
91 194.2 156.0 234.0









#69
o-Xylene
Concen: 19.585 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion:106 Resp: 4955
Ion Ratio Lower Upper
106 100
91 203.3 103.6 310.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#70
Styrene
Concen: 19.021 ug/l
RT: 12.056 min Scan# 1679
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:104 Resp: 79885
Ion Ratio Lower Upper
104 100
78 44.8 37.0 55.4
103 54.8 44.2 66.2





#71

Bromoform

Concen: 17.002 ug/l

RT: 12.221 min Scan# 1706

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion:	173	Resp:	13969
Ion Ratio	Lower	Upper	
173	100		
175	48.6	25.1	75.3
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.440 min Scan# 1906


Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion:	152	Resp:	84144
Ion Ratio	Lower	Upper	
152	100		
115	56.1	28.2	84.7
150	163.5	0.0	345.6





#73

Isopropylbenzene

Concen: 20.823 ug/l

RT: 12.343 min Scan# 1

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :

VY0320SBS01

Tgt Ion:105 Resp: 13017

Ion Ratio Lower Upper

105 100





120 26.2 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024

Supervised By :Semsettin Yesilyurt 03/21/2024



#74

N-ethyl acetate

Concen: 17.579 ug/l

RT: 12.154 min Scan# 1695

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion: 43 Resp: 19382




Ion Ratio Lower Upper

43 100

70 52.1 41.4 62.2

55 31.5 23.3 34.9

61 28.1 22.6 33.8





#75

1,1,2,2-Tetrachloroethane

Concen: 19.298 ug/l

RT: 12.593 min Scan# 1767

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion: 83	Resp: 20710
Ion Ratio	Lower Upper
83 100	
131 9.3	5.0 14.9
85 64.4	32.0 96.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#76

1,2,3-Trichloropropane

Concen: 19.135 ug/l m




RT: 12.642 min Scan# 1775

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion: 75	Resp: 15298
Ion Ratio	Lower Upper
75 100	
77 0.0	0.0 0.0





#77

Bromobenzene

Concen: 19.486 ug/l

RT: 12.623 min Scan# 1772

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion:	156	Resp:	29500
Ion Ratio	Lower	Upper	
156	100		
77	169.2	87.1	261.3
158	96.7	48.6	145.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#78

n-propylbenzene

Concen: 21.514 ug/l




RT: 12.684 min Scan# 1782


Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion:	91	Resp:	161375
Ion Ratio	Lower	Upper	
91	100		
120	22.8	11.5	34.4









#79
2-Chlorotoluene
Concen: 20.813 ug/l
RT: 12.770 min Scan# 11
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion: 91 Resp: 87400
Ion Ratio Lower Upper
91 100
126 35.9 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#80
1,3,5-Trimethylbenzene
Concen: 20.917 ug/l
RT: 12.824 min Scan# 1805
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:105 Resp: 104759
Ion Ratio Lower Upper
105 100
120 50.6 25.1 75.2









#81
trans-1,4-Dichloro-2-butene
Concen: 17.983 ug/l
RT: 12.392 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion: 75 Resp: 625
Ion Ratio Lower Upper
75 100
53 81.4 66.8 100.2
89 41.8 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#82
4-Chlorotoluene
Concen: 20.928 ug/l
RT: 12.867 min Scan# 1812
Delta R.T. 0.012 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 91 Resp: 89896
Ion Ratio Lower Upper
91 100
126 35.1 17.6 52.9





#83

tert-Butylbenzene

Concen: 20.788 ug/l

RT: 13.087 min Scan# 1847

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBSD01

Tgt Ion:	119	Resp:	9142
Ion Ratio	Lower	Upper	
119	100		
91	61.3	31.3	93.8
134	27.0	13.4	40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#84

1,2,4-Trimethylbenzene

Concen: 20.516 ug/l




RT: 13.135 min Scan# 1856


Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40


Tgt Ion:	105	Resp:	101440
Ion Ratio	Lower	Upper	
105	100		
120	47.0	23.2	69.6





#85
sec-Butylbenzene
Concen: 21.347 ug/l
RT: 13.263 min Scan# 1876
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40




Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01



Tgt Ion:105 Resp: 14353
Ion Ratio Lower Upper
105 100
134 20.1 10.1 30.1




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#86
p-Isopropyltoluene
Concen: 21.061 ug/l
RT: 13.379 min Scan# 1896
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:119 Resp: 115925
Ion Ratio Lower Upper
119 100
134 25.7 13.1 39.3
91 21.7 10.7 32.1









#87
1,3-Dichlorobenzene
Concen: 19.621 ug/l
RT: 13.379 min Scan# 1895
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBS01

Tgt Ion:146 Resp: 5962
Ion Ratio Lower Upper
146 100
111 37.7 19.0 57.0
148 63.2 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#88
1,4-Dichlorobenzene
Concen: 19.842 ug/l
RT: 13.458 min Scan# 1909
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion:146 Resp: 58617
Ion Ratio Lower Upper
146 100
111 36.9 18.4 55.4
148 63.1 32.4 97.2





#89

n-Butylbenzene

Concen: 21.185 ug/l

RT: 13.708 min Scan# 1949

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion: 91 Resp: 10638
 Ion Ratio Lower Upper
 91 100
 92 51.7 26.6 79.7
 134 24.8 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
 Supervised By :Semsettin Yesilyurt 03/21/2024



#90

Hexachloroethane

Concen: 20.996 ug/l




RT: 13.971 min Scan# 1993


Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion:117 Resp: 23169
 Ion Ratio Lower Upper
 117 100
 201 72.7 37.4 112.2









#91
1,2-Dichlorobenzene
Concen: 19.207 ug/l
RT: 13.751 min Scan# 1957
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Instrument :
MSVOA_Y
ClientSampleId :
VY0320SBSD01

Tgt Ion:146 Resp: 4936
Ion Ratio Lower Upper
146 100
111 39.3 19.9 59.7
148 65.1 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 16.420 ug/l
RT: 14.367 min Scan# 2058
Delta R.T. 0.006 min
Lab File: VY017722.D
Acq: 20 Mar 2024 10:40

Tgt Ion: 75 Resp: 2478
Ion Ratio Lower Upper
75 100
155 101.2 48.6 145.8
157 114.9 62.7 188.2





#93

1,2,4-Trichlorobenzene

Concen: 18.712 ug/l

RT: 15.019 min Scan# 2164

Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion:180	Resp:	25404
Ion Ratio	Lower	Upper
180	100	
182	92.9	47.9 143.8
145	28.9	14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#94

Hexachlorobutadiene

Concen: 20.542 ug/l




RT: 15.123 min Scan# 2182


Delta R.T. 0.006 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion:225	Resp:	17219
Ion Ratio	Lower	Upper
225	100	
223	62.1	30.4 91.2
227	65.2	31.9 95.9





#95

Naphthalene

Concen: 16.920 ug/l

RT: 15.251 min Scan# 2203

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Instrument :

MSVOA_Y

ClientSampleId :





VY0320SBS01

Tgt Ion:128	Resp:	37119
Ion Ratio	Lower	Upper
128	100	
127	13.1	10.6 16.0
129	10.9	8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/21/2024
Supervised By :Semsettin Yesilyurt 03/21/2024



#96

1,2,3-Trichlorobenzene

Concen: 18.273 ug/l

RT: 15.440 min Scan# 2234

Delta R.T. 0.012 min

Lab File: VY017722.D

Acq: 20 Mar 2024 10:40

Tgt Ion:180	Resp:	21282
Ion Ratio	Lower	Upper
180	100	
182	92.1	47.6 142.9
145	29.3	15.4 46.1

