



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032024\
Data File : VY017718.D
Acq On : 20 Mar 2024 08:31
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y022324S.M
Title : SW846 8260
Last Update : Sat Feb 24 04:41:20 2024



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1743

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.7	6128	PASS
75	95	30	60	50.5	16522	PASS
95	95	100	100	100.0	32741	PASS
96	95	5	9	6.8	2219	PASS
173	174	0.00	2	1.3	412	PASS
174	95	50	100	94.6	30965	PASS
175	174	5	9	7.1	2192	PASS
176	174	95	101	95.5	29563	PASS
177	176	5	9	6.3	1851	PASS