

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032123\
 Data File : VY013033.D
 Acq On : 21 Mar 2023 17:32
 Operator : KP/MD
 Sample : 02020-11RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-2RE

Quant Time: Mar 22 04:22:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	4375	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	6336	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	5801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2161	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	3574	82.278	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 164.560%#	
35) Dibromofluoromethane	7.716	113	2860	74.795	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 149.600%#	
50) Toluene-d8	10.179	98	7452	53.309	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.620%	
62) 4-Bromofluorobenzene	12.483	95	1725	36.560	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 73.120%	
Target Compounds						
16) Acetone	3.936	43	827	48.505	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032123\
Data File : VY013033.D
Acq On : 21 Mar 2023 17:32
Operator : KP/MD
Sample : 02020-11RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-2RE

Quant Time: Mar 22 04:22:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

