

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032125\
 Data File : VY021591.D
 Acq On : 21 Mar 2025 11:30
 Operator : SY/MD
 Sample : VY0321SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0321SBL01

Quant Time: Mar 21 23:09:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	271938	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	492142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	435285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	172430	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	161399	56.154	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.300%
35) Dibromofluoromethane	7.634	113	164595	50.881	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.760%
50) Toluene-d8	10.109	98	599353	48.933	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.860%
62) 4-Bromofluorobenzene	12.408	95	182435	43.787	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032125\
Data File : VY021591.D
Acq On : 21 Mar 2025 11:30
Operator : SY/MD
Sample : VY0321SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0321SBL01

Quant Time: Mar 21 23:09:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

