

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032125\
 Data File : VY021603.D
 Acq On : 21 Mar 2025 16:36
 Operator : SY/MD
 Sample : Q1606-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT24743

Quant Time: Mar 21 23:12:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.725	168	88	50.000	ug/l	# 0.01
34) 1,4-Difluorobenzene	0.000	114	0m	50.000	ug/l	-8.62
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.42
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.35

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	58 - 134	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	30 - 143	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032125\
Data File : VY021603.D
Acq On : 21 Mar 2025 16:36
Operator : SY/MD
Sample : Q1606-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT24743

Quant Time: Mar 21 23:12:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

