

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032221\
 Data File : VY004156.D
 Acq On : 22 Mar 2021 12:11
 Operator : SY/MD
 Sample : M1746-09
 Misc : 7.32g/5mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Mar 23 01:08:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	183906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	294737	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	263535	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	128525	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	98401	54.22	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.44%	
35) Dibromofluoromethane	7.728	113	88082	53.00	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.00%	
50) Toluene-d8	10.185	98	346706	53.21	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.42%	
62) 4-Bromofluorobenzene	12.483	95	121847	50.63	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.26%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.979	59	3028	18.45	ug/l	# 86
16) Acetone	3.960	43	29704	67.81	ug/l	# 89
18) Methyl Acetate	4.485	43	2037	1.88	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032221\
Data File : VY004156.D
Acq On : 22 Mar 2021 12:11
Operator : SY/MD
Sample : M1746-09
Misc : 7.32g/5mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Mar 23 01:08:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

