

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032221\
 Data File : VY004157.D
 Acq On : 22 Mar 2021 12:35
 Operator : SY/MD
 Sample : M1746-10
 Misc : 6.86g/5mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-4

Quant Time: Mar 23 01:08:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	189255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	323605	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	307032	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	149088	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	111613	59.76	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.52%
35) Dibromofluoromethane	7.728	113	99470	54.51	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.02%
50) Toluene-d8	10.185	98	383575	53.62	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.24%
62) 4-Bromofluorobenzene	12.483	95	142115	53.79	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.58%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032221\
Data File : VY004157.D
Acq On : 22 Mar 2021 12:35
Operator : SY/MD
Sample : M1746-10
Misc : 6.86g/5mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-4

Quant Time: Mar 23 01:08:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

