

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032224\
 Data File : VY017742.D
 Acq On : 22 Mar 2024 12:23
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Mar 23 03:45:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 03:33:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	127904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	242937	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	202977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	75752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	68004	61.030	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.060%	
35) Dibromofluoromethane	7.734	113	78129	53.801	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.600%	
50) Toluene-d8	10.197	98	279845	50.352	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 100.700%	
62) 4-Bromofluorobenzene	12.495	95	73487	44.033	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 88.060%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032224\
Data File : VY017742.D
Acq On : 22 Mar 2024 12:23
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Mar 23 03:45:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 03:33:45 2024
Response via : Initial Calibration

