

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032224\
 Data File : VY017736.D
 Acq On : 22 Mar 2024 09:44
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
 Supervised By :Mahesh Dadoda 03/25/2024

Quant Time: Mar 23 03:40:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 03:33:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	142441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	230710	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	192814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	82801	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	5154	4.153	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	8.300%#		
35) Dibromofluoromethane	7.734	113	5881	4.264	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	8.520%#		
50) Toluene-d8	10.197	98	21511	4.076	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	8.160%#		
62) 4-Bromofluorobenzene	12.495	95	6923	4.368	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	8.740%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.912	85	6937	5.641	ug/l	98
3) Chloromethane	2.125	50	8564	5.376	ug/l	89
4) Vinyl Chloride	2.265	62	10334	5.365	ug/l	91
5) Bromomethane	2.668	94	8224	5.468	ug/l	96
6) Chloroethane	2.808	64	7096	5.512	ug/l	96
7) Trichlorofluoromethane	3.143	101	11467	5.648	ug/l	100
8) Diethyl Ether	3.540	74	3198	4.778	ug/l	87
9) 1,1,2-Trichlorotrifluo...	3.918	101	7999	5.535	ug/l	98
10) Methyl Iodide	4.107	142	8465	4.792	ug/l	97
11) Tert butyl alcohol	4.972	59	2108	25.539	ug/l #	88
12) 1,1-Dichloroethene	3.893	96	7033	5.174	ug/l	93
13) Acrolein	3.747	56	2161	21.720	ug/l	91
14) Allyl chloride	4.503	41	8499	4.613	ug/l	96
15) Acrylonitrile	5.186	53	5746	20.723	ug/l	97
16) Acetone	3.972	43	4423	20.378	ug/l #	82
17) Carbon Disulfide	4.216	76	24141	5.307	ug/l	96
18) Methyl Acetate	4.503	43	2254	4.125	ug/l #	46
19) Methyl tert-butyl Ether	5.247	73	12052	4.112	ug/l	99
20) Methylene Chloride	4.741	84	9158	5.802	ug/l	93
21) trans-1,2-Dichloroethene	5.240	96	7988	5.154	ug/l	93
22) Diisopropyl ether	6.137	45	17121	4.268	ug/l #	88
23) Vinyl Acetate	6.082	43	44864	19.042	ug/l	100
24) 1,1-Dichloroethane	6.039	63	13161	4.992	ug/l	98
25) 2-Butanone	7.008	43	6132	24.071	ug/l	93
26) 2,2-Dichloropropane	6.996	77	11360	5.069	ug/l	97
27) cis-1,2-Dichloroethene	7.008	96	8543	4.873	ug/l	95
28) Bromochloromethane	7.356	49	4697	4.595	ug/l	89
29) Tetrahydrofuran	7.368	42	3704	20.075	ug/l	95
30) Chloroform	7.521	83	14091	5.155	ug/l	94
31) Cyclohexane	7.801	56	13350	5.439	ug/l #	79
32) 1,1,1-Trichloroethane	7.716	97	11946	5.018	ug/l	94
36) 1,1-Dichloropropene	7.935	75	10610	4.980	ug/l	94
37) Ethyl Acetate	7.088	43	3155	3.854	ug/l #	75
38) Carbon Tetrachloride	7.917	117	10996	4.929	ug/l	96
39) Methylcyclohexane	9.197	83	12691	4.640	ug/l #	87
40) Benzene	8.173	78	30260	4.745	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032224\
 Data File : VY017736.D
 Acq On : 22 Mar 2024 09:44
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
 Supervised By :Mahesh Dadoda 03/25/2024

Quant Time: Mar 23 03:40:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 03:33:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.332	41	1775	4.389	ug/l #	79
42) 1,2-Dichloroethane	8.252	62	6923	4.532	ug/l	97
43) Isopropyl Acetate	8.289	43	5544	4.088	ug/l	98
44) Trichloroethene	8.959	130	8369	5.017	ug/l	90
45) 1,2-Dichloropropane	9.234	63	6813	4.689	ug/l	95
46) Dibromomethane	9.325	93	3604	4.451	ug/l	96
47) Bromodichloromethane	9.508	83	10254	4.952	ug/l #	98
48) Methyl methacrylate	9.313	41	2408	4.386	ug/l	95
49) 1,4-Dioxane	9.313	88	659	83.641	ug/l #	91
51) 4-Methyl-2-Pentanone	10.087	43	13187	20.448	ug/l	93
52) Toluene	10.258	92	17835	4.526	ug/l	99
53) t-1,3-Dichloropropene	10.477	75	7462	4.121	ug/l	93
54) cis-1,3-Dichloropropene	9.947	75	9585	4.333	ug/l	92
55) 1,1,2-Trichloroethane	10.660	97	4583	4.424	ug/l	93
56) Ethyl methacrylate	10.526	69	4234	4.128	ug/l	92
57) 1,3-Dichloropropane	10.807	76	7828	4.387	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.801	63	11169	24.806	ug/l	100
59) 2-Hexanone	10.849	43	8152	22.293	ug/l	91
60) Dibromochloromethane	10.996	129	6445	4.652	ug/l	100
61) 1,2-Dibromoethane	11.105	107	4336	4.459	ug/l	99
64) Tetrachloroethene	10.733	164	9451	5.300	ug/l	96
65) Chlorobenzene	11.532	112	20195	4.866	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.605	131	6456	4.658	ug/l	97
67) Ethyl Benzene	11.611	91	32692	4.586	ug/l	96
68) m/p-Xylenes	11.715	106	24501	8.860	ug/l	99
69) o-Xylene	12.044	106	10862	4.302	ug/l	98
70) Styrene	12.062	104	17040	4.055	ug/l	97
71) Bromoform	12.221	173	3301	4.522	ug/l #	92
73) Isopropylbenzene	12.343	105	29210	4.672	ug/l	99
74) N-amyl acetate	12.160	43	3858	5.178	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.599	83	4181	4.343	ug/l	99
76) 1,2,3-Trichloropropane	12.648	75	3691m	4.704	ug/l	
77) Bromobenzene	12.623	156	6816	4.788	ug/l	98
78) n-propylbenzene	12.684	91	35133	4.642	ug/l	99
79) 2-Chlorotoluene	12.770	91	19822	4.776	ug/l	100
80) 1,3,5-Trimethylbenzene	12.831	105	21821	4.458	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.392	75	1369	4.137	ug/l	98
82) 4-Chlorotoluene	12.867	91	20052	4.709	ug/l	97
83) tert-Butylbenzene	13.093	119	20634	4.659	ug/l	100
84) 1,2,4-Trimethylbenzene	13.135	105	20664	4.307	ug/l	97
85) sec-Butylbenzene	13.270	105	30966	4.665	ug/l	98
86) p-Isopropyltoluene	13.385	119	24557	4.455	ug/l	98
87) 1,3-Dichlorobenzene	13.385	146	14308	4.960	ug/l	99
88) 1,4-Dichlorobenzene	13.465	146	13903	5.021	ug/l	93
89) n-Butylbenzene	13.708	91	22856	4.584	ug/l	97
90) Hexachloroethane	13.977	117	6060	5.391	ug/l	99
91) 1,2-Dichlorobenzene	13.751	146	11967	5.008	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.373	75	392	3.000	ug/l	83
93) 1,2,4-Trichlorobenzene	15.025	180	4939	4.267	ug/l	98
94) Hexachlorobutadiene	15.129	225	4066	5.472	ug/l	95
95) Naphthalene	15.257	128	5977	5.255	ug/l	99
96) 1,2,3-Trichlorobenzene	15.446	180	4139	4.331	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032224\
Data File : VY017736.D
Acq On : 22 Mar 2024 09:44
Operator : SY/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024

Quant Time: Mar 23 03:40:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 03:33:45 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.807 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:168 Resp: 14244
Ion Ratio Lower Upper
168 100
99 48.9 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#2
Dichlorodifluoromethane
Concen: 5.641 ug/l
RT: 1.912 min Scan# 15
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 85 Resp: 6937
Ion Ratio Lower Upper
85 100
87 30.8 16.0 48.0









#3
Chloromethane
Concen: 5.376 ug/l
RT: 2.125 min Scan# 50
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument : MSVOA_Y
ClientSampleId : VSTDICC005

Tgt Ion: 50 Resp: 8564
Ion Ratio Lower Upper
50 100
52 38.3 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#4
Vinyl Chloride
Concen: 5.365 ug/l
RT: 2.265 min Scan# 73
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44


Tgt Ion: 62 Resp: 10334
Ion Ratio Lower Upper
62 100
64 27.8 26.1 39.1





#5
Bromomethane
Concen: 5.468 ug/l
RT: 2.668 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 94 Resp: 8224
Ion Ratio Lower Upper
94 100
96 88.4 74.1 111.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024




#6
Chloroethane
Concen: 5.512 ug/l
RT: 2.808 min Scan# 162
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44



Tgt Ion: 64 Resp: 7096
Ion Ratio Lower Upper
64 100
66 30.4 26.1 39.1









#7
Trichlorofluoromethane
Concen: 5.648 ug/l
RT: 3.143 min Scan# 218
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:101 Resp: 1146
Ion Ratio Lower Upper
101 100
103 61.5 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#8
Diethyl Ether
Concen: 4.778 ug/l
RT: 3.540 min Scan# 282
Delta R.T. -0.012 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 74 Resp: 3198
Ion Ratio Lower Upper
74 100
45 94.0 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 5.535 ug/l
RT: 3.918 min Scan# 346
Delta R.T. -0.012 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:101 Resp: 7995
Ion Ratio Lower Upper
101 100
85 44.5 34.5 51.7
151 75.9 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#10
Methyl Iodide
Concen: 4.792 ug/l
RT: 4.107 min Scan# 375
Delta R.T. -0.012 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion:142 Resp: 8465
Ion Ratio Lower Upper
142 100
127 43.4 32.8 49.2
141 14.6 11.4 17.2









#11
Tert butyl alcohol
Concen: 25.539 ug/l
RT: 4.972 min Scan# 518
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 59 Resp: 2108
Ion Ratio Lower Upper
59 100
57 0.0 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#12
1,1-Dichloroethene
Concen: 5.174 ug/l
RT: 3.893 min Scan# 340
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 96 Resp: 7033
Ion Ratio Lower Upper
96 100
61 150.9 114.6 171.8
98 70.4 50.8 76.2









#13
Acrolein
Concen: 21.720 ug/l
RT: 3.747 min Scan# 317
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 56 Resp: 216
Ion Ratio Lower Upper
56 100
55 76.8 55.7 83.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#14
Allyl chloride
Concen: 4.613 ug/l
RT: 4.503 min Scan# 440
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44


Tgt Ion: 41 Resp: 8499
Ion Ratio Lower Upper
41 100
39 77.0 58.3 87.5
76 47.3 39.0 58.6





#15
Acrylonitrile
Concen: 20.723 ug/l
RT: 5.186 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 53 Resp: 5740
Ion Ratio Lower Upper
53 100
52 82.4 64.7 97.1
51 32.3 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024




#16
Acetone
Concen: 20.378 ug/l
RT: 3.972 min Scan# 353
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44



Tgt Ion: 43 Resp: 4423
Ion Ratio Lower Upper
43 100
58 28.5 31.7 47.5#





#17

Carbon Disulfide

Concen: 5.307 ug/l

RT: 4.216 min Scan# 394

Delta R.T. -0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 2414

Ion Ratio Lower Upper

76 100





78 9.8 6.8 10.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#18

Methyl Acetate

Concen: 4.125 ug/l

RT: 4.503 min Scan# 440

Delta R.T. -0.006 min

Lab File: VY017736.D




Acq: 22 Mar 2024 09:44


Tgt Ion: 43 Resp: 2254

Ion Ratio Lower Upper

43 100

74 0.0 23.3 34.9#









#19
Methyl tert-butyl Ether
Concen: 4.112 ug/l
RT: 5.247 min Scan# 501
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 73 Resp: 12052
Ion Ratio Lower Upper
73 100
57 22.2 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#20
Methylene Chloride
Concen: 5.802 ug/l
RT: 4.741 min Scan# 479
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44


Tgt Ion: 84 Resp: 9158
Ion Ratio Lower Upper
84 100
49 99.6 84.4 126.6
51 27.6 26.0 39.0
86 55.8 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 5.154 ug/l
RT: 5.240 min Scan# 501
Delta R.T. -0.012 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 96 Resp: 7988
Ion Ratio Lower Upper
96 100
61 133.2 97.5 146.3
98 63.3 50.5 75.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024




#22
Diisopropyl ether
Concen: 4.268 ug/l
RT: 6.137 min Scan# 708
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Tgt Ion: 45 Resp: 17121
Ion Ratio Lower Upper
45 100
43 64.7 42.2 63.2#
87 27.9 25.0 37.6
59 15.2 10.5 15.7





#23
 Vinyl Acetate
 Concen: 19.042 ug/l
 RT: 6.082 min Scan# 699
 Delta R.T. -0.006 min
 Lab File: VY017736.D
 Acq: 22 Mar 2024 09:44




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005




Tgt Ion: 43 Resp: 44864
 Ion Ratio Lower Upper
 43 100
 86 13.2 10.6 16.0

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 03/25/2024
 Supervised By :Mahesh Dadoda 03/25/2024




#24
 1,1-Dichloroethane
 Concen: 4.992 ug/l
 RT: 6.039 min Scan# 692
 Delta R.T. -0.006 min
 Lab File: VY017736.D
 Acq: 22 Mar 2024 09:44



Tgt Ion: 63 Resp: 13161
 Ion Ratio Lower Upper
 63 100
 98 9.1 4.0 12.2
 100 4.8 2.6 8.0









#25
2-Butanone
Concen: 24.071 ug/l
RT: 7.008 min Scan# 851
Delta R.T. -0.001 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 6132
Ion Ratio Lower Upper
43 100
72 38.3 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#26
2,2-Dichloropropane
Concen: 5.069 ug/l
RT: 6.996 min Scan# 849
Delta R.T. -0.012 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 77 Resp: 11360
Ion Ratio Lower Upper
77 100
97 22.7 12.0 36.1





#27
cis-1,2-Dichloroethene
Concen: 4.873 ug/l
RT: 7.008 min Scan# 851
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion: 96 Resp: 854
Ion Ratio Lower Upper
96 100
61 136.7 0.0 256.2
98 65.3 0.0 130.0

Manual Integrations




APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#28
Bromochloromethane
Concen: 4.595 ug/l
RT: 7.356 min Scan# 908
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 49 Resp: 4697
Ion Ratio Lower Upper
49 100
129 0.7 0.0 4.6
130 80.3 72.7 109.1





#29

Tetrahydrofuran

Concen: 20.075 ug/l

RT: 7.368 min Scan# 910

Delta R.T. 0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion: 42	Resp: 3704
Ion Ratio	Lower Upper
42 100	
72 51.6	43.0 64.4
71 45.5	40.8 61.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#30

Chloroform

Concen: 5.155 ug/l




RT: 7.521 min Scan# 935


Delta R.T. -0.012 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44


Tgt Ion: 83	Resp: 14091
Ion Ratio	Lower Upper
83 100	
85 60.5	52.2 78.4





#31
Cyclohexane
Concen: 5.439 ug/l
RT: 7.801 min Scan# 982
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 56 Resp: 13350
Ion Ratio Lower Upper
56 100
69 57.8 28.7 43.1
84 111.7 77.7 116.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#32
1,1,1-Trichloroethane
Concen: 5.018 ug/l
RT: 7.716 min Scan# 967
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 97 Resp: 11946
Ion Ratio Lower Upper
97 100
99 65.4 50.3 75.5
61 46.5 32.6 48.8





#33

1,2-Dichloroethane-d4

Concen: 4.153 ug/l

RT: 8.155 min Scan# 1109

Delta R.T. -0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 65 Resp: 5154

Ion Ratio Lower Upper

65 100





67 52.4 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.703 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44




Tgt Ion:114 Resp: 230710


Ion Ratio Lower Upper

114 100

63 18.9 0.0 35.0

88 14.5 0.0 27.2









#35
Dibromofluoromethane
Concen: 4.264 ug/l
RT: 7.734 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:113 Resp: 588
Ion Ratio Lower Upper
113 100
111 102.6 82.2 123.4
192 19.7 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#36
1,1-Dichloropropene
Concen: 4.980 ug/l
RT: 7.935 min Scan# 1003
Delta R.T. -0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 75 Resp: 10610
Ion Ratio Lower Upper
75 100
110 33.0 19.0 57.0
77 30.4 25.3 37.9





#37

Ethyl Acetate

Concen: 3.854 ug/l

RT: 7.088 min Scan# 80

Delta R.T. -0.012 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005





Tgt Ion: 43 Resp: 315

Ion	Ratio	Lower	Upper
43	100		
61	0.0	12.6	19.0
70	8.0	7.8	11.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#38

Carbon Tetrachloride

Concen: 4.929 ug/l

RT: 7.917 min Scan# 1000




Delta R.T. -0.006 min


Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:117 Resp: 10996

Ion	Ratio	Lower	Upper
117	100		
119	89.6	75.8	113.6
121	31.0	24.0	36.0





#39

Methylcyclohexane

Concen: 4.640 ug/l

RT: 9.197 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion: 83	Resp: 1269
Ion Ratio	Lower Upper
83 100	
55 78.0	51.5 77.3
98 45.5	40.5 60.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#40

Benzene

Concen: 4.745 ug/l




RT: 8.173 min Scan# 1042


Delta R.T. -0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44


Tgt Ion: 78	Resp: 30260
Ion Ratio	Lower Upper
78 100	
77 21.0	19.5 29.3





#41
Methacrylonitrile
Concen: 4.389 ug/l
RT: 7.332 min Scan# 904
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 41 Resp: 1775
Ion Ratio Lower Upper
41 100
39 49.7 48.0 72.0
67 69.5 75.1 112.7
52 24.2 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#42
1,2-Dichloroethane
Concen: 4.532 ug/l
RT: 8.252 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 62 Resp: 6923
Ion Ratio Lower Upper
62 100
98 10.0 0.0 22.0





#43

Isopropyl Acetate

Concen: 4.088 ug/l

RT: 8.289 min Scan# 1061

Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion: 43	Resp: 554		
Ion Ratio	Lower	Upper	
43	100		
61	34.5	27.9	41.9
87	14.8	13.1	19.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#44

Trichloroethene

Concen: 5.017 ug/l




RT: 8.959 min Scan# 1171


Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:130	Resp:	8369	
Ion	Ratio	Lower	Upper
130	100		
95	84.2	0.0	188.0





#45

1,2-Dichloropropane

Concen: 4.689 ug/l

RT: 9.234 min Scan# 1216

Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 63 Resp: 681

Ion Ratio Lower Upper

63 100





65 28.8 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#46

Dibromomethane

Concen: 4.451 ug/l

RT: 9.325 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44




Tgt Ion: 93 Resp: 3604


Ion Ratio Lower Upper

93 100

95 88.3 66.2 99.4

174 99.7 80.9 121.3









#47
Bromodichloromethane
Concen: 4.952 ug/l
RT: 9.508 min Scan# 1261
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 10254
Ion Ratio Lower Upper
83 100
85 63.1 50.1 75.1
127 6.0 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#48
Methyl methacrylate
Concen: 4.386 ug/l
RT: 9.313 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 41 Resp: 2408
Ion Ratio Lower Upper
41 100
69 101.3 84.1 126.1
39 60.5 43.6 65.4





#49

1,4-Dioxane

Concen: 83.641 ug/l

RT: 9.313 min Scan# 1229

Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 88 Resp: 659

Ion Ratio Lower Upper

88 100

43 22.3 0.0 0.0





58 54.6 49.3 73.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#50

Toluene-d8

Concen: 4.076 ug/l

RT: 10.197 min Scan# 1374

Delta R.T. -0.000 min

Lab File: VY017736.D




Acq: 22 Mar 2024 09:44


Tgt Ion: 98 Resp: 21511

Ion Ratio Lower Upper

98 100

100 63.3 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 20.448 ug/l
RT: 10.087 min Scan# 1355
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 1318
Ion Ratio Lower Upper
43 100
58 40.2 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#52
Toluene
Concen: 4.526 ug/l
RT: 10.258 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 92 Resp: 17835
Ion Ratio Lower Upper
92 100
91 175.0 139.0 208.6









#53
t-1,3-Dichloropropene
Concen: 4.121 ug/l
RT: 10.477 min Scan# 1420
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 75 Resp: 7462
Ion Ratio Lower Upper
75 100
77 26.0 24.1 36.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#54
cis-1,3-Dichloropropene
Concen: 4.333 ug/l
RT: 9.947 min Scan# 1333
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44


Tgt Ion: 75 Resp: 9585
Ion Ratio Lower Upper
75 100
77 34.4 25.4 38.0
39 43.4 29.8 44.6





#55
1,1,2-Trichloroethane
Concen: 4.424 ug/l
RT: 10.660 min Scan# 1449
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion:	97	Resp:	458
Ion	Ratio	Lower	Upper
97	100		
83	94.4	67.5	101.3
85	51.1	42.9	64.3
99	60.5	46.6	70.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#56
Ethyl methacrylate
Concen: 4.128 ug/l
RT: 10.526 min Scan# 1428
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion:	69	Resp:	4234
Ion	Ratio	Lower	Upper
69	100		
41	57.9	42.3	63.5
39	33.8	22.9	34.3





#57

1,3-Dichloropropane

Concen: 4.387 ug/l

RT: 10.807 min Scan# 1473

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 7828

Ion Ratio Lower Upper

76 100





78 33.1 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#58

2-Chloroethyl Vinyl ether

Concen: 24.806 ug/l

RT: 9.801 min Scan# 1309

Delta R.T. 0.006 min

Lab File: VY017736.D




Acq: 22 Mar 2024 09:44


Tgt Ion: 63 Resp: 11169

Ion Ratio Lower Upper

63 100

106 29.8 23.7 35.5









#59
2-Hexanone
Concen: 22.293 ug/l
RT: 10.849 min Scan# 1481
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 815
Ion Ratio Lower Upper
43 100
58 55.9 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#60
Dibromochloromethane
Concen: 4.652 ug/l
RT: 10.996 min Scan# 1505
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44


Tgt Ion:129 Resp: 6445
Ion Ratio Lower Upper
129 100
127 77.6 39.0 117.0





#61
1,2-Dibromoethane
Concen: 4.459 ug/l
RT: 11.105 min Scan# 1523
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:107 Resp: 4330
Ion Ratio Lower Upper
107 100
109 93.5 75.9 113.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024




#62
4-Bromofluorobenzene
Concen: 4.368 ug/l
RT: 12.495 min Scan# 1751
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44



Tgt Ion: 95 Resp: 6923
Ion Ratio Lower Upper
95 100
174 92.5 0.0 175.6
176 87.0 0.0 171.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.508 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion:117 Resp: 192814

Ion Ratio Lower Upper

117 100

82 51.8 42.4 63.6





119 32.8 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#64

Tetrachloroethene

Concen: 5.300 ug/l

RT: 10.733 min Scan# 1462

Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:164 Resp: 9451




Ion Ratio Lower Upper


164 100

166 122.2 102.6 154.0

129 87.2 72.0 108.0

131 83.2 70.0 105.0









#65
Chlorobenzene
Concen: 4.866 ug/l
RT: 11.532 min Scan# 1153
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:112 Resp: 2019
Ion Ratio Lower Upper
112 100
114 33.2 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 4.658 ug/l
RT: 11.605 min Scan# 1605
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion:131 Resp: 6456
Ion Ratio Lower Upper
131 100
133 94.1 47.8 143.3
119 72.4 34.1 102.2





#67

Ethyl Benzene

Concen: 4.586 ug/l

RT: 11.611 min Scan# 1606

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 32692

Ion Ratio Lower Upper

91 100





106 32.9 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#68

m/p-Xylenes

Concen: 8.860 ug/l

RT: 11.715 min Scan# 1623

Delta R.T. -0.000 min

Lab File: VY017736.D




Acq: 22 Mar 2024 09:44


Tgt Ion:106 Resp: 24501

Ion Ratio Lower Upper

106 100

91 193.4 156.0 234.0









#69
o-Xylene
Concen: 4.302 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:106 Resp: 10862
Ion Ratio Lower Upper
106 100
91 210.3 103.6 310.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#70
Styrene
Concen: 4.055 ug/l
RT: 12.062 min Scan# 1680
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44


Tgt Ion:104 Resp: 17040
Ion Ratio Lower Upper
104 100
78 47.7 37.0 55.4
103 58.2 44.2 66.2





#71
Bromoform
Concen: 4.522 ug/l
RT: 12.221 min Scan# 1706
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion:173 Resp: 330
Ion Ratio Lower Upper
173 100
175 45.0 25.1 75.3
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.440 min Scan# 1906
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44



Tgt Ion:152 Resp: 82801
Ion Ratio Lower Upper
152 100
115 57.0 28.2 84.7
150 157.9 0.0 345.6





#73

Isopropylbenzene

Concen: 4.672 ug/l

RT: 12.343 min Scan# 1726

Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	105	Resp:	29210
Ion Ratio	Lower	Upper	
105	100		
120	26.4	13.4	40.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#74

N-ethyl acetate

Concen: 5.178 ug/l




RT: 12.160 min Scan# 1696


Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44


Tgt Ion:	43	Resp:	3858
Ion Ratio	Lower	Upper	
43	100		
70	50.2	41.4	62.2
55	28.0	23.3	34.9
61	26.5	22.6	33.8





#75
1,1,2,2-Tetrachloroethane
Concen: 4.343 ug/l
RT: 12.599 min Scan# 1767
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion: 83 Resp: 418
Ion Ratio Lower Upper
83 100
131 10.9 5.0 14.9
85 64.5 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024




#76
1,2,3-Trichloropropane
Concen: 4.704 ug/l m
RT: 12.648 min Scan# 1776
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44



Tgt Ion: 75 Resp: 3691
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 4.788 ug/l

RT: 12.623 min Scan# 1772

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	156	Resp:	6810
Ion Ratio	Lower	Upper	
156	100		
77	172.1	87.1	261.3
158	100.3	48.6	145.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#78

n-propylbenzene

Concen: 4.642 ug/l




RT: 12.684 min Scan# 1782


Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:	91	Resp:	35133
Ion Ratio	Lower	Upper	
91	100		
120	23.4	11.5	34.4









#79
2-Chlorotoluene
Concen: 4.776 ug/l
RT: 12.770 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 91 Resp: 1982
Ion Ratio Lower Upper
91 100
126 35.6 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#80
1,3,5-Trimethylbenzene
Concen: 4.458 ug/l
RT: 12.831 min Scan# 1806
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion:105 Resp: 21821
Ion Ratio Lower Upper
105 100
120 49.7 25.1 75.2









#81
trans-1,4-Dichloro-2-butene
Concen: 4.137 ug/l
RT: 12.392 min Scan# 1734
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 75 Resp: 1369
Ion Ratio Lower Upper
75 100
53 86.8 66.8 100.2
89 44.8 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#82
4-Chlorotoluene
Concen: 4.709 ug/l
RT: 12.867 min Scan# 1812
Delta R.T. -0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 91 Resp: 20052
Ion Ratio Lower Upper
91 100
126 37.2 17.6 52.9





#83

tert-Butylbenzene

Concen: 4.659 ug/l

RT: 13.093 min Scan# 1849

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	119	Resp:	20634
Ion Ratio	Lower	Upper	
119	100		
91	62.6	31.3	93.8
134	27.2	13.4	40.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#84

1,2,4-Trimethylbenzene

Concen: 4.307 ug/l




RT: 13.135 min Scan# 1856


Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:	105	Resp:	20664
Ion Ratio	Lower	Upper	
105	100		
120	48.6	23.2	69.6









#85
 sec-Butylbenzene
 Concen: 4.665 ug/l
 RT: 13.270 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VY017736.D
 Acq: 22 Mar 2024 09:44

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005

Tgt Ion:105 Resp: 3096
 Ion Ratio Lower Upper
 105 100
 134 20.8 10.1 30.1




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
 Supervised By :Mahesh Dadoda 03/25/2024



#86
 p-Isopropyltoluene
 Concen: 4.455 ug/l
 RT: 13.385 min Scan# 1897
 Delta R.T. 0.006 min
 Lab File: VY017736.D
 Acq: 22 Mar 2024 09:44

Tgt Ion:119 Resp: 24557
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.1 39.3
 91 22.7 10.7 32.1





#87

1,3-Dichlorobenzene

Concen: 4.960 ug/l

RT: 13.385 min Scan# 1896

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion:146 Resp: 14308

Ion Ratio Lower Upper

146 100

111 36.8 19.0 57.0





148 63.6 31.7 95.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024

Supervised By :Mahesh Dadoda 03/25/2024



#88

1,4-Dichlorobenzene

Concen: 5.021 ug/l

RT: 13.465 min Scan# 1910

Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44




Tgt Ion:146 Resp: 13903


Ion Ratio Lower Upper

146 100

111 45.3 18.4 55.4

148 66.8 32.4 97.2





#89

n-Butylbenzene

Concen: 4.584 ug/l

RT: 13.708 min Scan# 1950

Delta R.T. -0.000 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	91	Resp:	2285
Ion	Ratio	Lower	Upper
91	100		
92	50.7	26.6	79.7
134	24.5	12.9	38.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#90

Hexachloroethane

Concen: 5.391 ug/l




RT: 13.977 min Scan# 1994


Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:	117	Resp:	6060
Ion	Ratio	Lower	Upper
117	100		
201	73.9	37.4	112.2









#91
1,2-Dichlorobenzene
Concen: 5.008 ug/l
RT: 13.751 min Scan# 1957
Delta R.T. 0.000 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:146 Resp: 1196
Ion Ratio Lower Upper
146 100
111 40.0 19.9 59.7
148 63.2 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 3.000 ug/l
RT: 14.373 min Scan# 2059
Delta R.T. 0.006 min
Lab File: VY017736.D
Acq: 22 Mar 2024 09:44

Tgt Ion: 75 Resp: 392
Ion Ratio Lower Upper
75 100
155 111.0 48.6 145.8
157 146.7 62.7 188.2





#93

1,2,4-Trichlorobenzene

Concen: 4.267 ug/l

RT: 15.025 min Scan# 2166

Delta R.T. 0.012 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:180	Resp:	4939
Ion Ratio	Lower	Upper
180	100	
182	98.3	47.9 143.8
145	28.6	14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#94

Hexachlorobutadiene

Concen: 5.472 ug/l




RT: 15.129 min Scan# 2183


Delta R.T. 0.006 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:225	Resp:	4066
Ion Ratio	Lower	Upper
225	100	
223	64.2	30.4 91.2
227	68.6	31.9 95.9





#95

Naphthalene

Concen: 5.255 ug/l

RT: 15.257 min Scan# 21

Delta R.T. 0.012 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:128	Resp: 597
Ion Ratio	Lower Upper
128 100	
127 12.9	10.6 16.0
129 11.5	8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 03/25/2024
Supervised By :Mahesh Dadoda 03/25/2024



#96

1,2,3-Trichlorobenzene

Concen: 4.331 ug/l

RT: 15.446 min Scan# 2235

Delta R.T. 0.012 min

Lab File: VY017736.D

Acq: 22 Mar 2024 09:44

Tgt Ion:180	Resp: 4139
Ion Ratio	Lower Upper
180 100	
182 97.9	47.6 142.9
145 32.1	15.4 46.1

