

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032321\
 Data File : VY004171.D
 Acq On : 23 Mar 2021 12:16
 Operator : SY/MD
 Sample : M1765-01
 Misc : 5.83g/5mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RBR-07

Quant Time: Mar 24 01:18:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	204136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	343144	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	326576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	163342	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	113405	56.29	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.58%
35) Dibromofluoromethane	7.734	113	105332	54.43	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.86%
50) Toluene-d8	10.185	98	408392	53.83	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.66%
62) 4-Bromofluorobenzene	12.483	95	149764	53.45	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.90%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032321\
Data File : VY004171.D
Acq On : 23 Mar 2021 12:16
Operator : SY/MD
Sample : M1765-01
Misc : 5.83g/5mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RBR-07

Quant Time: Mar 24 01:18:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

