

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032321\
 Data File : VY004173.D
 Acq On : 23 Mar 2021 13:02
 Operator : SY/MD
 Sample : M1780-01
 Misc : 5.21g/5mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TT-41-RW6-IDWSO-20210322

Quant Time: Mar 24 01:19:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	206077	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	348570	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	327218	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	158294	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	117715	57.88	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.76%
35) Dibromofluoromethane	7.728	113	106139	54.00	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.00%
50) Toluene-d8	10.184	98	417238	54.14	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.28%
62) 4-Bromofluorobenzene	12.483	95	149894	52.67	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.34%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032321\
Data File : VY004173.D
Acq On : 23 Mar 2021 13:02
Operator : SY/MD
Sample : M1780-01
Misc : 5.21g/5mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TT-41-RW6-IDWSO-20210322

Quant Time: Mar 24 01:19:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

