



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032321\
Data File : VY004164.D
Acq On : 23 Mar 2021 08:47
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Title : SW846 8260
Last Update : Mon Mar 08 12:28:55 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	16408	PASS
75	95	30	60	52.9	44888	PASS
95	95	100	100	100.0	84877	PASS
96	95	5	9	6.8	5805	PASS
173	174	0.00	2	1.1	739	PASS
174	95	50	100	82.4	69979	PASS
175	174	5	9	7.6	5317	PASS
176	174	95	101	96.3	67371	PASS
177	176	5	9	6.9	4662	PASS