

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032323\
 Data File : VY013068.D
 Acq On : 23 Mar 2023 18:22
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 24 05:06:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.493	0.441	10.5	87	0.00
3 P	Chloromethane	0.589	0.460	21.9	78	0.00
4 C	Vinyl Chloride	0.540	0.472	12.6#	82	0.00
5 T	Bromomethane	0.328	0.310	5.5	90	-0.01
6 T	Chloroethane	0.324	0.303	6.5	87	-0.01
7 T	Trichlorofluoromethane	0.897	0.735	18.1	78	0.00
8 T	Diethyl Ether	0.314	0.311	1.0	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.540	0.505	6.5	88	0.00
10 T	Methyl Iodide	0.703	0.652	7.3	81	0.00
11 T	Tert butyl alcohol	0.070	0.062	11.4	102	-0.01
12 CM	1,1-Dichloroethene	0.515	0.477	7.4#	86	-0.01
13 T	Acrolein	0.087	0.061	29.9#	66	0.00
14 T	Allyl chloride	0.951	0.902	5.2	87	0.00
15 T	Acrylonitrile	0.166	0.180	-8.4	99	-0.01
16 T	Acetone	0.195	0.193	1.0	85	0.00
17 T	Carbon Disulfide	1.753	1.535	12.4	82	0.00
18 T	Methyl Acetate	0.580	0.617	-6.4	97	0.00
19 T	Methyl tert-butyl Ether	1.330	1.411	-6.1	95	0.00
20 T	Methylene Chloride	0.714	0.621	13.0	94	0.00
21 T	trans-1,2-Dichloroethene	0.560	0.551	1.6	91	-0.01
22 T	Diisopropyl ether	1.727	1.858	-7.6	97	-0.01
23 T	Vinyl Acetate	1.073	1.188	-10.7	97	0.00
24 P	1,1-Dichloroethane	0.984	1.006	-2.2	94	0.00
25 T	2-Butanone	0.241	0.267	-10.8	97	0.00
26 T	2,2-Dichloropropane	0.874	0.843	3.5	91	-0.01
27 T	cis-1,2-Dichloroethene	0.600	0.616	-2.7	93	0.00
28 T	Bromochloromethane	0.360	0.364	-1.1	91	0.00
29 T	Tetrahydrofuran	0.146	0.168	-15.1	102	0.00
30 C	Chloroform	0.972	1.009	-3.8#	95	0.00
31 T	Cyclohexane	0.951	0.885	6.9	88	0.00
32 T	1,1,1-Trichloroethane	0.879	0.882	-0.3	92	0.00
33 S	1,2-Dichloroethane-d4	0.496	0.551	-11.1	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.302	0.370	-22.5	111	0.00
36 T	1,1-Dichloropropene	0.506	0.510	-0.8	91	0.00
37 T	Ethyl Acetate	0.319	0.353	-10.7	103	0.00
38 T	Carbon Tetrachloride	0.522	0.538	-3.1	93	0.00
39 T	Methylcyclohexane	0.601	0.602	-0.2	88	0.00
40 TM	Benzene	1.430	1.467	-2.6	92	0.00
41 T	Methacrylonitrile	0.172	0.188	-9.3	97	0.00
42 TM	1,2-Dichloroethane	0.430	0.454	-5.6	94	0.00
43 T	Isopropyl Acetate	0.569	0.635	-11.6	98	0.00
44 TM	Trichloroethene	0.377	0.381	-1.1	90	0.00
45 C	1,2-Dichloropropane	0.361	0.382	-5.8#	97	0.00
46 T	Dibromomethane	0.206	0.223	-8.3	96	0.00
47 T	Bromodichloromethane	0.485	0.518	-6.8	96	0.00
48 T	Methyl methacrylate	0.262	0.305	-16.4	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032323\
 Data File : VY013068.D
 Acq On : 23 Mar 2023 18:22
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 24 05:06:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	112	0.00
50 S	Toluene-d8	1.103	1.142	-3.5	95	0.00
51 T	4-Methyl-2-Pentanone	0.308	0.364	-18.2	101	0.00
52 CM	Toluene	0.863	0.906	-5.0#	93	0.00
53 T	t-1,3-Dichloropropene	0.509	0.554	-8.8	96	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.606	-5.2	93	0.00
55 T	1,1,2-Trichloroethane	0.287	0.308	-7.3	97	0.00
56 T	Ethyl methacrylate	0.382	0.436	-14.1	96	0.00
57 T	1,3-Dichloropropane	0.489	0.534	-9.2	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.227	-12.9	96	0.00
59 T	2-Hexanone	0.224	0.263	-17.4	99	0.00
60 T	Dibromochloromethane	0.349	0.378	-8.3	97	0.00
61 T	1,2-Dibromoethane	0.276	0.298	-8.0	96	0.00
62 S	4-Bromofluorobenzene	0.372	0.446	-19.9	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.364	0.383	-5.2	97	0.00
65 PM	Chlorobenzene	1.049	1.066	-1.6	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.405	-4.4	96	0.00
67 C	Ethyl Benzene	1.845	1.907	-3.4#	93	0.00
68 T	m/p-Xylenes	0.710	0.738	-3.9	93	0.00
69 T	o-Xylene	0.662	0.686	-3.6	93	0.00
70 T	Styrene	1.105	1.194	-8.1	95	0.00
71 P	Bromoform	0.253	0.272	-7.5	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.544	3.425	3.4	93	0.00
74 T	N-amyl acetate	1.115	1.178	-5.7	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.797	0.817	-2.5	101	0.00
76 T	1,2,3-Trichloropropane	0.597	0.557	6.7	91	0.00
77 T	Bromobenzene	0.861	0.822	4.5	94	0.00
78 T	n-propylbenzene	4.369	4.263	2.4	93	0.00
79 T	2-Chlorotoluene	2.428	2.362	2.7	94	0.00
80 T	1,3,5-Trimethylbenzene	2.924	2.879	1.5	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.288	0.286	0.7	96	0.00
82 T	4-Chlorotoluene	2.513	2.458	2.2	95	0.00
83 T	tert-Butylbenzene	2.524	2.528	-0.2	95	0.00
84 T	1,2,4-Trimethylbenzene	2.860	2.830	1.0	94	0.00
85 T	sec-Butylbenzene	3.851	3.785	1.7	95	0.00
86 T	p-Isopropyltoluene	3.147	3.089	1.8	94	0.00
87 T	1,3-Dichlorobenzene	1.660	1.619	2.5	95	0.00
88 T	1,4-Dichlorobenzene	1.694	1.618	4.5	94	0.00
89 T	n-Butylbenzene	3.013	2.922	3.0	93	0.00
90 T	Hexachloroethane	0.655	0.620	5.3	95	0.00
91 T	1,2-Dichlorobenzene	1.501	1.483	1.2	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.129	0.132	-2.3	97	0.00
93 T	1,2,4-Trichlorobenzene	0.924	0.843	8.8	87	0.00
94 T	Hexachlorobutadiene	0.564	0.495	12.2	89	0.00
95 T	Naphthalene	1.852	1.798	2.9	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032323\
Data File : VY013068.D
Acq On : 23 Mar 2023 18:22
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Mar 24 05:06:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.822	0.751	8.6	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6