

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032323\
 Data File : VY013062.D
 Acq On : 23 Mar 2023 15:08
 Operator : KP/MD
 Sample : 02028-01RE
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-032123RE

Quant Time: Mar 24 05:04:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	28640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	48151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	43449	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	18938	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	22292	78.394	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	156.780%
35) Dibromofluoromethane	7.710	113	19173	65.979	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	131.960%
50) Toluene-d8	10.179	98	56606	53.284	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.560%
62) 4-Bromofluorobenzene	12.483	95	16145	45.027	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.060%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032323\
Data File : VY013062.D
Acq On : 23 Mar 2023 15:08
Operator : KP/MD
Sample : 02028-01RE
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-032123RE

Quant Time: Mar 24 05:04:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY013062.D
Acq: 23 Mar 2023 15:08




Instrument :
MSVOA_Y
ClientSampleId :
TR-05-032123RE

Tgt Ion:168 Resp: 28640
Ion Ratio Lower Upper
168 100
99 49.9 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 78.394 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.007 min
Lab File: VY013062.D
Acq: 23 Mar 2023 15:08

Tgt Ion: 65 Resp: 22292
Ion Ratio Lower Upper
65 100
67 52.6 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.006 min

Lab File: VY013062.D

Acq: 23 Mar 2023 15:08

Instrument :

MSVOA_Y

ClientSampleId :

TR-05-032123RE





Tgt Ion:114 Resp: 48151

Ion Ratio Lower Upper

114 100

63 20.2 0.0 37.4

88 14.1 0.0 28.6



#35

Dibromofluoromethane

Concen: 65.979 ug/l

RT: 7.710 min Scan# 966

Delta R.T. -0.006 min

Lab File: VY013062.D

Acq: 23 Mar 2023 15:08




Tgt Ion:113 Resp: 19173

Ion Ratio Lower Upper

113 100

111 101.9 82.2 123.2

192 20.3 16.8 25.2








#50
Toluene-d8
Concen: 53.284 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY013062.D
Acq: 23 Mar 2023 15:08




Instrument :
MSVOA_Y
ClientSampleId :
TR-05-032123RE


Tgt Ion: 98 Resp: 56606
Ion Ratio Lower Upper
98 100
100 63.0 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 45.027 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY013062.D
Acq: 23 Mar 2023 15:08

Tgt Ion: 95 Resp: 16145
Ion Ratio Lower Upper
95 100
174 89.9 0.0 176.6
176 87.0 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. -0.000 min
Lab File: VY013062.D
Acq: 23 Mar 2023 15:08

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-032123RE

Tgt Ion:	117	Resp:	43449
Ion	Ratio	Lower	Upper
117	100		
82	53.1	45.4	68.0
119	33.2	25.6	38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. -0.006 min
Lab File: VY013062.D
Acq: 23 Mar 2023 15:08

Tgt Ion:	152	Resp:	18938
Ion	Ratio	Lower	Upper
152	100		
115	53.2	27.1	81.3
150	153.9	0.0	347.8

