

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032421\
 Data File : VY004182.D
 Acq On : 24 Mar 2021 12:52
 Operator : SY/MD
 Sample : M1786-02
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-1A

Quant Time: Mar 25 01:18:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	195450	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	330060	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	318054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	162070	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	112296	58.22	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.44%
35) Dibromofluoromethane	7.728	113	101019	54.27	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.54%
50) Toluene-d8	10.185	98	388803	53.28	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.56%
62) 4-Bromofluorobenzene	12.483	95	147801	54.84	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.68%
Target Compounds						
16) Acetone	3.966	43	6346	13.63	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032421\
Data File : VY004182.D
Acq On : 24 Mar 2021 12:52
Operator : SY/MD
Sample : M1786-02
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-1A

Quant Time: Mar 25 01:18:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

