

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032421\
 Data File : VY004185.D
 Acq On : 24 Mar 2021 14:02
 Operator : SY/MD
 Sample : M1730-01
 Misc : 5.08g/5mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-8A

Quant Time: Mar 25 01:19:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	184171	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	301417	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	280371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	139860	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	96497	53.09	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.18%	
35) Dibromofluoromethane	7.734	113	90660	53.34	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.68%	
50) Toluene-d8	10.185	98	356894	53.56	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.12%	
62) 4-Bromofluorobenzene	12.483	95	128575	52.24	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.48%	
Target Compounds						Qvalue
16) Acetone	3.960	43	3937	8.97	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032421\
Data File : VY004185.D
Acq On : 24 Mar 2021 14:02
Operator : SY/MD
Sample : M1730-01
Misc : 5.08g/5mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-8A

Quant Time: Mar 25 01:19:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

